

REFERENCES

1. Jarvis K.E., Gray A.L., Houk R.S.; *Handbook of inductively coupled plasma mass spectrometry*; Blackie, pp.18-22, 1992.
2. Willard H.H., Merritt L.L., Dean J.A., Settle F.A.; *Instrumental methods of analysis*; 7th edition, Wadsworth, p.266, 1988.
3. Houk R.S.; *Mass spectrometry of inductively coupled plasmas*; Anal.Chem., **58**, pp.97A-105A, 1986.
4. Date A.R., Gray A.L., *Applications of inductively coupled plasma mass spectrometry*; Blackie Glasgow, pp.1-42, 1989.
5. Olivares J.A., Houk R.S.; *Suppression of analyte signal by various concomitant salts in inductively coupled plasma mass spectrometry*; Anal.Chem., **58**, pp.20-25, 1986.
6. Douglas D.J., French J.B.; *Gas dynamics of the inductively coupled plasma mass spectrometry interface*; J.Anal.Atom.Spectrom., **3**, pp.743-747, 1988.
7. Gray A.L.; *Mass spectrometry with an inductively coupled plasma as an ion source: the influence on ultratrace analysis of background and matrix response*; Spectrochim.Acta, **41B**, pp.151-167, 1986.
8. Douglas D.J., French J.B.; *An improved interface for inductively coupled plasma-mass spectrometry (ICP-MS)*; Spectrochim.Acta, **41B**, pp.197-204, 1986.
9. Gray A.L.; *Influence of load coil geometry on oxide and doubly charged ion response in inductively coupled plasma source mass spectrometry*; J.Anal.Atom.Spectrom., **1**, pp.247-249, 1986.
10. Houk R.S., Svec H.J., Fassel V.A.; *Mass spectrometric evidence for suprathermal ionization in an inductively coupled argon plasma*; Appl.Spectrosc., **35**, pp.380-384, 1981.
11. Douglas D.J.; *Fundamental aspects of ICP-MS*; In *ICPs in Analytical Atomic Spectrometry*; 2nd edition, editors Montaser A., Golightly D.W., VCH Publishers, New York, 1991.
12. Hutton R.C., Eaton A.N.; *Role of aerosol water vapour loading in inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **2**, pp.595-598, 1987.
13. Zhu G., Browner R.F.; *Study of the influence of water vapour loading and interface pressure in ICP-MS*; J.Anal.Atom.Spectrom., **3**, pp.781-789, 1988.
14. Olivares J.A., Houk R.S.; *Ion sampling for inductively coupled plasma mass spectrometry*; Anal.Chem., **57**, pp.2674-2679, 1985.

15. Lam J.W.H., Horlick G.; *Effects of sampler-skimmer separation in inductively coupled plasma-mass spectrometry*; Spectrochim.Acta, **45B**, pp.1327-1338, 1990.
16. Crain J.S., Smith F.G., Houk R.S.; *Mass spectrometric measurement of ionization temperature in an inductively coupled plasma*; Spectrochim.Acta, **45B**, pp.249-259, 1990.
17. Wilson D.A., Vickers G.H., Hieftje G.M.; *Ionization temperatures in the inductively coupled plasma determined by mass spectrometry*; Appl.Spectrosc., **41**, pp.875- 880, 1987.
18. Olivares J.A., Houk R.S.; *Ion sampling for inductively coupled plasma mass spectrometry*; Anal.Chem., **57**, pp.2674-2679, 1985.
19. Gillson G.R., Douglas D.J., Fulford J.E., Halligan K.W., Tanner S.D.; *Nonspectroscopic interelement interferences in inductively coupled plasma mass spectrometry*; Anal.Chem., **60**, pp.1472-1474, 1988.
20. Houk R.S., Fassel V.A., Flesch G.D., Svec H.J., Gray A.L., Taylor C.E.; *Inductively coupled argon plasma as an ion source for mass spectrometric determination of trace elements*; Anal.Chem., **52**, pp.2283-2289, 1980.
21. Gregoire D.C.; *Influence of instrument parameters on nonspectroscopic interferences in inductively coupled plasma-mass spectrometry*; Appl.Spectrosc., **41(5)**, pp.897-903, 1987.
22. Thompson J.J., Houk R.S.; *A study of internal standardisation in inductively coupled plasma-mass spectrometry*; Appl.Spectrosc., **41(5)**, pp.801-806, 1987.
23. Vickers G.H., Ross B.S., Hieftje G.M.; *Reduction of mass-dependent interferences in inductively coupled plasma-mass spectrometry by using flow-injection analysis*; Appl.Spectrosc., **43**, pp.1330-1333, 1989.
24. Tan S.H., Horlick G.; *Matrix-effect observations in inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **2**, pp.745-763, 1987.
25. Beauchemin D., McLaren J.W., Berman S.S.; *Study of the effects on concomitant elements in inductively coupled plasma mass spectrometry*; Spectrochim. Acta, **42B**, pp.467-490, 1987.
26. Crain J.S., Houk R.S., Smith F.G.; *Matrix interferences in inductively coupled plasma-mass spectrometry: some effects of skimmer orifice diameter and ion lens voltages*; Spectrochim. Acta, **43B**, pp.1355-1364, 1988.
27. Wang J., Shen W.-L., Sheppard B.S., Evans E.H., Caruso J.A., Fricke F.L.; *Effect of ion lens tuning and flow injection on non-spectroscopic matrix interferences in inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **5**, pp.445- 449, 1990.
28. Kawaguchi H., Tanaka T., Mizuike A.; *Continuum background in ICP-MS*; Spectrochim.Acta, **43B**, pp.955-962, 1988.

29. Horlick G., Tan S.H., Vaughan M.A., Rose C.A.; *The effect of plasma operating parameters on analyte signals in inductively coupled plasma-mass spectrometry*; Spectrochim.Acta, **40B**, pp.1555-1572, 1985.
30. Vaughan M.A., Horlick G., Tan S.H., *Effect of operating parameters on analyte signals in inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **2**, pp.765-772, 1987.
31. Zhu G., Browner R.F.; *Investigation of experimental parameters with a quadrupole ICP/MS*; Appl.Spectrosc., **41(3)**, pp.349-359, 1987.
32. Geerling R., Hattendorf B., Schmidt K.P., Krengel-Rothensee K., *Determination of trace element concentrations in water samples using the spectromass 2000*; Spectro Report, **98**, pp.1-8, 1998.
33. Long S.E., Brown R.M.; *Optimisation in inductively coupled plasma mass spectrometry*; Analyst, **111**, pp.901-906, 1986.
34. Gray A.L., Williams J.G.; *System optimisation and the effect on polyatomic, oxide and doubly charged ion response of a commercial inductively coupled plasma mass spectrometry instrument*; J.Anal.Atom.Spectrom., **2**, pp.599-606, 1987.
35. Date A.R., Gray A.L.; *Plasma source mass spectrometry using an inductively coupled plasma and a high resolution quadrupole mass filter*; Analyst (London), **106**, pp.1255-1267, 1981.
36. Date A.R., Gray A.L.; *Development progress in plasma source mass spectrometry*; Analyst (London), **108**, pp.159-165, 1983.
37. Gray A.L., Date A.R.; *Inductively coupled plasma source mass spectrometry using continuum flow ion extraction*; Analyst (London), **108**, pp.1033-1050, 1983.
38. Date A.R., Gray A.L.; *Progress in plasma source mass spectrometry*; Spectrochim.Acta, **38B**, pp.29-37, 1983.
39. Douglas D.J., Quan E.S.K., Smith R.G.; *Elemental analysis with an atmospheric pressure plasma (MIP, ICP) / quadrupole mass spectrometer system*; Spectrochim.Acta, **38B**, pp.39-48, 1983.
40. Doherty W.; *An internal standardization procedure for the determination of yttrium and the rare earth elements in geological materials by inductively coupled plasma- mass spectrometry*; Spectrochim.Acta, **44B**, pp.263-280, 1989.
41. Palmieri M.D., Fritz J.S., Thompson J.J., Houk R.S.; *Separation of trace rare earths and other metals from uranium by liquid-liquid extraction with quantitation by inductively-coupled plasma / mass spectrometry*; Anal.Chim.Acta, **184**, pp.187-196, 1986.

42. Doherty W., Van der Voet A.; *The application of inductively coupled plasma mass spectrometry to the determination of rare earth elements in geological materials*; Can.J.Spectrosc., **30**(6), pp.135-141, 1985.
43. McLaren J., Beauchemin D., Van der Voet T.; *ICP-MS activities in Canada*; Can.J.Spectrosc., **30**, pp.29A-32A, 1985.
44. McLaren J.W., Mykytiuk A.P., Willie S.N., Berman S.S.; *Determination of trace metals in seawater by inductively coupled plasma mass spectrometry with preconcentration on silica-immobilized 8-hydroxyquinoline*; Anal.Chem., **57**, pp.2907-2911, 1985.
45. Vandecasteele C., Nagels M., Vanhoe H., Dams R.; *Suppression of analyte signal in inductively-coupled plasma / mass spectrometry and the use of an internal standard*; Anal.Chim.Acta, **211**, pp.91-98, 1988.
46. Gregoire D.C.; *The effect of easily ionizable concomitant elements on non-spectroscopic interferences in inductively coupled plasma-mass spectrometry*; Spectrochim.Acta, **42B**, pp.895-907, 1987.
47. Hall G.E.M., Park C.J., Pelchat J.C.; *Determination of tungsten and molybdenum at low levels in geological materials by inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **2**, pp.189-196, 1987.
48. McLaren J.W., Beauchemin D., Berman S.S.; *Analysis of the marine sediment reference material PACS-1 by inductively coupled plasma mass spectrometry*; Spectrochim.Acta, **43B**, pp.413-420, 1988.
49. Chen X., Houk R.S.; *Polyatomic ions as internal standards for matrix corrections in inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **10**, pp.837-841, 1995.
50. Jarvis K.E., Gray A.L., Houk R.S.; *Handbook of inductively coupled plasma mass spectrometry*; Blackie, pp.341-347, 1992.
51. Gowing C.J.B., Potts P.J.; *Evaluation of a rapid technique for the determination of precious metals in geological samples based on a selective aqua regia leach*; Analyst, **116**, pp.773-779, 1991.
52. Juvonen R., Kallio E., Lakomaa T.; *Determination of precious metals in rocks by inductively coupled plasma mass spectrometry using nickel sulfide concentration. Comparison with other pre-treatment methods*; Analyst, **119**, pp.617-621, 1994.
53. Rentoul E., Smith H.; *Toxic materials in medical jurisprudence and toxicology*; 13th edition, Churchill Livingstone, pp.537-546, 1973.
54. Calmus Y., Poupon R.; *Foie et arsenic*; Gastroenterologie Clinique et Biologique, **6**, pp.933-941, 1982.

55. Fuortes L.; *Arsenic poisoning. Ongoing diagnostic and social problem*; Postgraduate Medicine, **83(1)**, pp.234-244, 1988.
56. Levin-Scherz J.K., Patrick J.D., Weber F.H., Garabedian C.; *Acute arsenic ingestion*; Annals of Emergency Medicine, **16(6)**, pp.702-704, 1987.
57. Gerhardt R.E., Hudson J.B., Rao R.N., Sobel R.E.; *Chronic renal insufficiency from cortical necrosis induced by arsenic poisoning*; Archives of Internal Medicine, **138**, pp.1267-1269, 1978.
58. Welter A., Michaux H., Blondel A.; *Lignes de mees dans un cas d'intoxication aiguë par l'arsenic*; Dermatologica, **165(5)**, pp.482-483, 1982.
59. Quatrehomme G., Ricq O., Lapalus P., Jacomet Y., Ollier A.; *Acute arsenic intoxication: forensic and toxicological aspects (an observation)*; J.For.Sciences, **37(4)**, pp.1163-1171, 1992.
60. Wojeck G.A., Nigg H.N., Braman R., Stamper J.M., Roussef R.; *Worker exposure to arsenic in Florida grapefruit spray operations*; Archives of Environmental Contamination and Toxicology, **11**, pp.661-667, 1982.
61. Fazekas I.G., Rengei B.; *Sur la teneur normale d'arsenic dans le cheveux, les poils axillaires, et les poils du pubis, selon sexes et ages*; Annales de Médecine Légale, **40**, pp.35-40, 1960.
62. Planques J., Brustier V., Bourbon P., Pitrt G., Broussy G.; *Contribution à l'étude de la répartition de l'arsenic dans l'organisme humain au cours d'une intoxication chronique collective*; Annales de Médecine Légale, **40**, pp.509-515, 1960.
63. Eckert W.G.; *Introduction to forensic sciences forensic toxicology*; The C.V. Mosby Company, p.99, 1980.
64. Smith H.; J.For.Medicine, **9(4)**, 1962.
65. Bagchi K.N.; Indian Med.Gazette, **72**, p.477, 1937.
66. Smales A.A., Pate E.D.; *The detection of sub-microgram quantities of arsenic by radioactivation. Part 3 The detection of arsenic in biological material*; Analyst, **77**, p.196, 1952.
67. Heydorn K.; *Environmental variation of arsenic levels in human blood determined by neutron activation analysis*; Clin.Chim.Acta, **28**, pp.349-357, 1970.
68. Espinoza E.O., Mann M-J., Bleasdell B., DeKorte S., Cox M.; *Toxic metals in selected traditional chinese medicinals*; J.For.Sciences, **41(3)**, pp.453-456, 1996.
69. Tay C.H., Seah C.S.; *Arsenic poisoning from anti-asthmatic herbal preparations*; Med.J.Australia, **2**, pp.424-428, 1975.
70. Wijesekera A.R.L., Henry K.D., Ranasinghe P.; *The detection and estimation of (a) arsenic in opium, and (b) strychnine in opium and heroin, as a means of identification of their respective sources*; For.Science Int., **36**, pp.193-209, 1988.

71. Sunshine I., *Methodology for analytical toxicology*; 2nd edition, C.C. Thomas, pp.30-33, 1978.
72. Curry A.S.; *Poison detection in human organs*; 2nd edition, C.C. Thomas, pp.159-163, 1969.
73. Kaye S.; *Handbook of emergency toxicology*; 3rd edition, C.C. Thomas, pp.48-57, 1973.
74. Sunshine I.; *CRC manual of analytical toxicology*; Chemical Rubber Company, pp.32-35, 1971.
75. Maes D., Pate B.D.; *The absorption of arsenic into single human head hairs*; J.For.Sciences, **22**(1), pp.89-94, 1977.
76. Heitkemper D.T., Kaine L.A., Jackson D.S., Wolnik K.A.; *Practical applications of element-specific detection by inductively coupled plasma atomic emission spectroscopy and inductively coupled plasma mass spectrometry to ion chromatography of foods*; J.Chromatogr., **671A**, pp.101-108, 1994.
77. Tanaka T., Hara K., Tanimoto A., Kasai K., Kita T., Tanaka N., Takayasu T.; *Determination of arsenic in blood and stomach contents by inductively coupled plasma / mass spectrometry (ICP/MS)*; For.Science Int., **81**, pp.43-50, 1996.
78. Kershisnik M.M., Kalamegham R., Ash K.O., Nixon D.E., Ashwood E.R., *Using ¹⁶O³⁵Cl to correct for chloride interference improves accuracy of urine arsenic determinations by inductively coupled plasma mass spectrometry*; Clin.Chem., **38**(11), pp.2197-2202, 1992.
79. Campbell M.J., Demesmay C., Ollé M., *Determination of total arsenic concentrations in biological matrices by inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **9**, pp.1379-1384, 1994.
80. Branch S., Ebdon L., Ford M., Foulkes M., O'Neill P., *Determination of arsenic in samples with high chloride content by inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **6**, pp.151-154, 1991.
81. Nixon D.E., Moyer T.P.; *Routine clinical determination of lead, arsenic, cadmium, and thallium in urine and whole blood by inductively coupled plasma mass spectrometry*; Spectrochim.Acta, **51B**, pp.13-25, 1996.
82. Goergen M.G., Murshak V.F., Roettger P., Murshak I., Edelman D.; *ICP-MS analysis of toxic characteristics leaching procedure (TCLP) extract: advantages and disadvantages*; At.Spectrosc., **13**(1), pp.11-18, 1992.
83. Jarvis K.E., Gray A.L., Houk R.S.; *Handbook of inductively coupled plasma mass spectrometry*; Blackie, pp.129-134, 1992.
84. Sakata K., Kawabata K.; *Reduction of fundamental polyatomic ions in inductively coupled plasma mass spectrometry*; Spectrochim.Acta, **49B**, pp.1027-1038, 1994.
85. Vaughan M.A., Horlick G.; *Oxide, hydroxide, and doubly charged analyte species in inductively coupled plasma / mass spectrometry*; Appl.Spectrosc., **40**(4), pp.434- 445, 1986.

86. Tan S.H., Horlick G.; *Background spectral features in inductively coupled plasma / mass spectrometry*; Appl.Spectrosc., **40(4)**, pp.445-460, 1986.
87. Prohaska T., Latkoczy C., Stinger G., Wenzel W.W., Blum W.E.; *HR-ICPMS determination of arsenic in environmental samples - spectral interferences and matrix effects*; Presented at ISEAC26, April 1996.
88. Branch S., Ebdon L., O'Neill P., *Determination of arsenic species in fish by directly coupled high-performance liquid chromatography-inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **9**, pp.33-37, 1994.
89. Anderson S.T.G., Robért R.V.D., Farrer H.N.; *Determination of total and leachable arsenic and selenium in soils by continuous hydride generation inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **9**, pp.1107-1110, 1994.
90. Ebdon L., Fisher A.S., Worsfold P.J., *Determination of arsenic, chromium, selenium and vanadium in biological samples by inductively coupled plasma mass spectrometry using on-line elimination of interference and preconcentration by flow injection*; J.Anal.Atom.Spectrom., **9**, pp.611-614, 1994.
91. Date A.R., Cheung Y.Y., Stuart M.E.; *The influence of polyatomic ion interferences in analysis by inductively coupled plasma mass spectrometry (ICP-MS)*; Spectrochim.Acta, **42B**, pp.3-20, 1987.
92. Amarasiriwardena D., Krushevsk A., Barnes R.M.; *Microwave-assisted vapor- phase nitric acid digestion of small biological samples for inductively coupled plasma spectrometry*; Appl.Spectrosc., **52(6)**, pp.900-907, 1998.
93. Sheppard B.S., Heitkemper D.T., Gaston C.M.; *Microwave digestion for the determination of arsenic, cadmium and lead in seafood products by inductively coupled plasma atomic emission and mass spectrometry*; Analyst, **119**, pp.1683- 1686, 1994.
94. Lásztity A., Krushevsk A., Kotrebai M., Barnes R.M., Amarasiriwardena D.; *Arsenic determination in environmental, biological and food samples by inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **10**, pp.505-510, 1995.
95. Steele T.W.; *Certificate of analysis - Platinum Ore - SARM 7 - Certified reference material*; pp.1-4, 1975.
96. Lenahan W.C., Murray-Smith R de L.; *Assay and analytical practice in the South African mining industry*; The Chamber of Mines of South Africa, The South African Institute of Mining and Metallurgy, Monograph series M6, 1986.

97. Hall G.E.M., Pelchat J.C.; *Analysis of geological materials for gold, platinum and palladium at low ppb levels by fire assay-ICP mass spectrometry*; Chem.Geol., **115**, pp.61-72, 1994.
98. Jackson S.E., Fryer B.J., Gosse W., Healey D.C., Longerich H.P., Strong D.F.; *Determination of the precious metals in geological materials by inductively coupled plasma-mass spectrometry (ICP-MS) with nickel sulphide fire-assay collection and tellurium coprecipitation*; Chem.Geol., **83**, pp.119-132, 1990.
99. Sun M., Jain J., Zhou M., Kerrich R.; *A procedural modification for enhanced recovery of precious metals (Au, PGE) following nickel sulphide fire assay and tellurium coprecipitation: applications for analysis of geological samples by inductively coupled plasma mass spectrometry*; Can.J.Appl.Spectrosc., **38(4)**, pp.103-108, 1993.
100. Perry B.J., Van Loon J.C., Speller D.V.; *Dry-chlorination inductively coupled plasma mass spectrometric method for the determination of platinum group elements in rocks*; J.Anal.Atom.Spectrom., **7**, pp.883-888, 1992.
101. Chen Z., Fryer B.J., Longerich H.P., Jackson S.E.; *Determination of the precious metals in milligram samples of sulfides and oxides using inductively coupled plasma mass spectrometry after ion exchange preconcentration*; J.Anal.Atom.Spectrom., **11**, pp.805-809, 1996.
102. Enzweiler J., Potts P.J., Jarvis K.E.; *Determination of platinum, palladium, ruthenium and iridium in geological samples by isotope dilution inductively coupled plasma mass spectrometry using a sodium peroxide fusion and tellurium coprecipitation*; Analyst, **120**, pp.1391-1396, 1995.
103. Godfrey J., McCurdy E.; *Investigation into the feasibility of ICP-MS as an alternative to fire assay measurements for gold and the platinum group elements*; Applications of plasma source mass spectrometry II; Holland G., Eaton A.N. (editors); pp.65-71, 1993.
104. Enzweiler J., Potts P.J.; *The separation of platinum, palladium and gold from silicate rocks by the anion exchange separation of chloro complexes after a sodium peroxide fusion: an investigation of low recoveries*; Talanta, **42**, pp.1411-1418, 1995.
105. Yi Y.V., Masuda A.; *Isotopic homogenization of iridium for high sensitivity determination by isotope dilution inductively coupled plasma spectrometry*; Anal.Sciences, **12**, pp.7-12, 1996.
106. Perry B.J., Barefoot R.R., Van Loon J.C.; *Inductively coupled plasma mass spectrometry for the determination of platinum group elements and gold*; Trends in Anal.Chem., **14(8)**, pp.388-397, 1995.

107. Qu Y.B.; *Recent developments in the determination of precious metals*; Analyst, **121**, pp.139-161, 1996.
108. Seronorm Trace Elements Urine; Lot 403125, 403125x, 403125y; Analytical values; Mat.no: 800106; Certificate of analysis.
109. Larsen E.H., Stürup S.; *Carbon-enhanced inductively coupled plasma mass spectrometric detection of arsenic and selenium and its application to arsenic speciation*; J.Anal.Atom.Spectrom., **9**, pp.1099-1105, 1994.
110. Wang C., Jeng S., Shieh F.; *Determination of arsenic in airborne particulate matter by inductively coupled plasma mass spectrometry*; J.Anal.Atom.Spectrom., **12**, pp.61-67, 1997.
111. Sakao S., Uchida H.; *Determination of trace elements in shellfish tissue samples by inductively coupled plasma mass spectrometry*; Anal.Chim.Acta, **382**, pp.215-223, 1999.

ADDENDUM A

AVERAGES OF THE INTENSITIES MEASURED OF THE ISOTOPES OF THE INTERNAL STANDARDS, THE PLATINUM GROUP ELEMENTS AND GOLD

Table 3.5: Averages of the measured intensities (counts s⁻¹) of the different isotopes of Ar, Sc, Y, La, Au and Ir.

Sample	³⁶ Ar	⁴⁵ Sc	⁸⁹ Y	¹³⁸ La	¹³⁹ La	¹⁹⁷ Au	¹⁹¹ Ir	¹⁹³ Ir
Blank in 1% v/v HCl	2.06x10 ⁷	9.83x10 ⁵	6.98x10 ⁵	4.58x10 ²	3.53x10 ⁵	1.16x10 ²	7.12x10 ¹	7.44x10 ¹
10 µg dm ⁻³ of element in 1% v/v HCl	2.01x10 ⁷	9.03x10 ⁵	6.45x10 ⁵	4.50x10 ²	3.36x10 ⁵	7.93x10 ²	6.76x10 ²	1.15x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.00x10 ⁷	8.81x10 ⁵	6.21x10 ⁵	4.60x10 ²	3.22x10 ⁵	3.52x10 ³	3.11x10 ³	5.03x10 ³
100 µg dm ⁻³ of element in 1% v/v HCl	2.05x10 ⁷	9.18x10 ⁵	6.33x10 ⁵	4.70x10 ²	3.28x10 ⁵	7.12x10 ³	6.28x10 ³	1.03x10 ⁴
150 µg dm ⁻³ of element in 1% v/v HCl	1.97x10 ⁷	9.22x10 ⁵	6.63x10 ⁵	4.24x10 ²	3.39x10 ⁵	1.10x10 ⁴	9.76x10 ³	1.60x10 ⁴
50 µg dm ⁻³ of element in 1% v/v HCl	2.05x10 ⁷	9.33x10 ⁵	6.65x10 ⁵	5.10x10 ²	3.36x10 ⁵	3.73x10 ³	3.28x10 ³	5.45x10 ³
50 µg dm ⁻³ of element in 0.35% v/v aqua regia	1.63x10 ⁷	8.31x10 ⁵	6.08x10 ⁵	4.29x10 ²	3.26x10 ⁵	3.49x10 ³	3.16x10 ³	5.11x10 ³
50 µg dm ⁻³ of element in 0.50% v/v aqua regia	1.70x10 ⁷	8.48x10 ⁵	6.09x10 ⁵	4.14x10 ²	3.29x10 ⁵	3.64x10 ³	3.25x10 ³	5.27x10 ³
50 µg dm ⁻³ of element in 1.00% v/v aqua regia	1.86x10 ⁷	8.73x10 ⁵	6.42x10 ⁵	4.49x10 ²	3.33x10 ⁵	3.58x10 ³	3.17x10 ³	5.25x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.02x10 ⁷	9.19x10 ⁵	6.61x10 ⁵	4.44x10 ²	3.36x10 ⁵	3.75x10 ³	3.32x10 ³	5.42x10 ³
50 µg dm ⁻³ of element in 1.50% v/v aqua regia	2.07x10 ⁷	9.33x10 ⁵	6.57x10 ⁵	4.51x10 ²	3.33x10 ⁵	3.67x10 ³	3.19x10 ³	5.32x10 ³
50 µg dm ⁻³ of element in 2.00% v/v aqua regia	2.23x10 ⁷	9.81x10 ⁵	6.90x10 ⁵	4.64x10 ²	3.48x10 ⁵	3.83x10 ³	3.33x10 ³	5.30x10 ³
50 µg dm ⁻³ of element in 2.50% v/v aqua regia	2.27x10 ⁷	9.75x10 ⁵	6.76x10 ⁵	4.80x10 ²	3.44x10 ⁵	3.71x10 ³	3.20x10 ³	5.45x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.01x10 ⁷	9.27x10 ⁵	6.61x10 ⁵	4.77x10 ²	3.35x10 ⁵	3.68x10 ³	3.17x10 ³	5.32x10 ³
Standard deviation of blank in 1% v/v HCl	1.05x10 ⁶	6.85x10 ⁴	5.90x10 ⁴	2.17x10 ¹	2.87x10 ⁴	1.22x10 ¹	6.33x10 ⁰	6.43x10 ⁰

Table 3.6: Averages of the measured intensities (counts s⁻¹) of the different isotopes of Pd and Pt.

Sample	¹⁰² Pd	¹⁰⁴ Pd	¹⁰⁵ Pd	¹⁰⁶ Pd	¹⁰⁸ Pd	¹¹⁰ Pd	¹⁹² Pt	¹⁹⁴ Pt	¹⁹⁵ Pt	¹⁹⁶ Pt	¹⁹⁸ Pt
Blank in 1% v/v HCl	8.43x10 ¹	1.07x10 ²	3.81x10 ³	6.76x10 ²	1.14x10 ²	1.01x10 ²	7.50x10 ¹	5.55x10 ¹	6.61x10 ¹	6.44x10 ¹	6.87x10 ¹
10 µg dm ⁻³ of element in 1% v/v HCl	1.82x10 ³	1.49x10 ³	4.37x10 ³	1.67x10 ³	1.10x10 ³	5.65x10 ²	9.67x10 ¹	4.34x10 ²	4.29x10 ²	3.62x10 ²	1.47x10 ²
50 µg dm ⁻³ of element in 1% v/v HCl	8.63x10 ³	7.28x10 ³	7.83x10 ³	6.13x10 ³	5.46x10 ³	2.46x10 ³	1.02x10 ²	1.74x10 ³	1.82x10 ³	1.40x10 ³	4.82x10 ²
100 µg dm ⁻³ of element in 1% v/v HCl	1.78x10 ⁴	1.49x10 ⁴	1.24x10 ⁴	1.20x10 ⁴	1.09x10 ⁴	5.08x10 ³	1.88x10 ²	3.58x10 ³	3.62x10 ³	2.68x10 ³	8.19x10 ²
150 µg dm ⁻³ of element in 1% v/v HCl	2.74x10 ⁴	2.31x10 ⁴	1.79x10 ⁴	1.81x10 ⁴	1.72x10 ⁴	7.76x10 ³	2.53x10 ²	5.59x10 ³	5.73x10 ³	4.23x10 ³	1.29x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	9.20x10 ³	7.83x10 ³	8.39x10 ³	6.45x10 ³	5.80x10 ³	2.56x10 ³	9.10x10 ¹	1.88x10 ³	1.88x10 ³	1.45x10 ³	4.73x10 ²
50 µg dm ⁻³ of element in 0.35% v/v aqua regia	8.44x10 ³	7.14x10 ³	7.60x10 ³	5.73x10 ³	5.21x10 ³	2.36x10 ³	1.09x10 ²	1.84x10 ³	1.83x10 ³	1.35x10 ³	4.45x10 ²
50 µg dm ⁻³ of element in 0.50% v/v aqua regia	8.71x10 ³	7.17x10 ³	7.64x10 ³	6.00x10 ³	5.19x10 ³	2.39x10 ³	1.33x10 ²	1.79x10 ³	1.92x10 ³	1.41x10 ³	4.80x10 ²
50 µg dm ⁻³ of element in 1.00% v/v aqua regia	8.77x10 ³	7.54x10 ³	8.15x10 ³	6.16x10 ³	5.53x10 ³	2.54x10 ³	1.09x10 ²	1.88x10 ³	1.88x10 ³	1.47x10 ³	4.91x10 ²
50 µg dm ⁻³ of element in 1% v/v HCl	9.23x10 ³	7.71x10 ³	8.09x10 ³	6.21x10 ³	5.69x10 ³	2.61x10 ³	1.07x10 ²	1.82x10 ³	1.89x10 ³	1.48x10 ³	4.94x10 ²
50 µg dm ⁻³ of element in 1.50% v/v aqua regia	9.09x10 ³	7.60x10 ³	8.34x10 ³	6.37x10 ³	5.76x10 ³	2.55x10 ³	1.24x10 ²	1.88x10 ³	1.94x10 ³	1.38x10 ³	4.69x10 ²
50 µg dm ⁻³ of element in 2.00% v/v aqua regia	9.64x10 ³	7.99x10 ³	8.75x10 ³	6.62x10 ³	5.74x10 ³	2.69x10 ³	1.16x10 ²	1.88x10 ³	1.93x10 ³	1.47x10 ³	4.92x10 ²
50 µg dm ⁻³ of element in 2.50% v/v aqua regia	9.14x10 ³	7.80x10 ³	8.55x10 ³	6.48x10 ³	5.74x10 ³	2.63x10 ³	1.26x10 ²	1.85x10 ³	1.83x10 ³	1.44x10 ³	4.86x10 ²
50 µg dm ⁻³ of element in 1% v/v HCl	9.17x10 ³	7.71x10 ³	8.25x10 ³	6.17x10 ³	5.67x10 ³	2.64x10 ³	1.34x10 ²	1.85x10 ³	1.96x10 ³	1.40x10 ³	4.88x10 ²
Standard deviation of blank in 1% v/v HCl	1.11x10 ¹	1.29x10 ¹	4.21x10 ²	6.58x10 ¹	1.29x10 ¹	2.00x10 ¹	1.01x10 ¹	2.06x10 ¹	1.54x10 ¹	1.54x10 ¹	1.33x10 ¹

Table 3.7: Averages of the measured intensities (counts s⁻¹) of the different isotopes of Rh and Ru.

Sample	¹⁰³ Rh	⁹⁶ Ru	⁹⁸ Ru	⁹⁹ Ru	¹⁰⁰ Ru	¹⁰¹ Ru	¹⁰² Ru	¹⁰⁴ Ru
Blank in 1% v/v HCl	1.24x10 ²	1.06x10 ²	1.60x10 ²	8.42x10 ¹	1.49x10 ²	8.99x10 ¹	9.21x10 ¹	1.05x10 ²
10 µg dm ⁻³ of element in 1% v/v HCl	4.68x10 ³	4.29x10 ²	2.88x10 ²	7.95x10 ²	8.15x10 ²	1.04x10 ³	1.81x10 ³	1.53x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.25x10 ⁴	1.49x10 ³	6.26x10 ²	3.36x10 ³	3.41x10 ³	4.69x10 ³	8.51x10 ³	7.24x10 ³
100 µg dm ⁻³ of element in 1% v/v HCl	4.65x10 ⁴	3.09x10 ³	1.21x10 ³	6.83x10 ³	6.96x10 ³	9.26x10 ³	1.75x10 ⁴	1.50x10 ⁴
150 µg dm ⁻³ of element in 1% v/v HCl	7.28x10 ⁴	4.53x10 ³	1.67x10 ³	1.05x10 ⁴	1.04x10 ⁴	1.42x10 ⁴	2.69x10 ⁴	2.26x10 ⁴
50 µg dm ⁻³ of element in 1% v/v HCl	2.41x10 ⁴	1.55x10 ³	6.80x10 ²	3.72x10 ³	3.74x10 ³	4.90x10 ³	9.34x10 ³	7.72x10 ³
50 µg dm ⁻³ of element in 0.35% v/v aqua regia	2.23x10 ⁴	1.55x10 ³	6.24x10 ²	3.38x10 ³	3.35x10 ³	4.48x10 ³	8.25x10 ³	7.18x10 ³
50 µg dm ⁻³ of element in 0.50% v/v aqua regia	2.26x10 ⁴	1.52x10 ³	6.31x10 ²	3.38x10 ³	3.43x10 ³	4.54x10 ³	8.42x10 ³	7.09x10 ³
50 µg dm ⁻³ of element in 1.00% v/v aqua regia	2.33x10 ⁴	1.53x10 ³	7.12x10 ²	3.49x10 ³	3.43x10 ³	4.72x10 ³	9.02x10 ³	7.51x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.41x10 ⁴	1.69x10 ³	7.18x10 ²	3.58x10 ³	3.55x10 ³	4.79x10 ³	9.12x10 ³	7.60x10 ³
50 µg dm ⁻³ of element in 1.50% v/v aqua regia	2.37x10 ⁴	1.62x10 ³	7.04x10 ²	3.57x10 ³	3.59x10 ³	4.75x10 ³	9.03x10 ³	7.55x10 ³
50 µg dm ⁻³ of element in 2.00% v/v aqua regia	2.47x10 ⁴	1.64x10 ³	7.54x10 ²	3.75x10 ³	3.64x10 ³	5.20x10 ³	9.50x10 ³	8.03x10 ³
50 µg dm ⁻³ of element in 2.50% v/v aqua regia	2.43x10 ⁴	1.65x10 ³	7.38x10 ²	3.58x10 ³	3.82x10 ³	4.97x10 ³	9.46x10 ³	7.75x10 ³
50 µg dm ⁻³ of element in 1% v/v HCl	2.39x10 ⁴	1.53x10 ³	6.77x10 ²	3.65x10 ³	3.64x10 ³	4.70x10 ³	9.18x10 ³	7.66x10 ³
Standard deviation of blank in 1% v/v HCl	3.67x10 ¹	6.44x10 ⁰	2.41x10 ¹	2.00x10 ¹	3.08x10 ¹	3.42x10 ¹	1.40x10 ¹	2.89x10 ¹

ADDENDUM B

CALIBRATION DATA FOR THE ISOTOPES OF THE PLATINUM GROUP ELEMENTS AND GOLD

Table 3.8: Calibration data for Au and Ir isotopes when no internal standard was used.

[Element] ($\mu\text{g dm}^{-3}$)	^{197}Au			^{191}Ir			^{193}Ir		
	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.16×10^2	1.32	-	7.12×10^1	1.18	-	7.44×10^1	1.03	-
10	7.93×10^2	10.65	6.52	6.76×10^2	10.59	5.92	1.15×10^3	11.21	12.09
50	3.52×10^3	48.21	-3.59	3.11×10^3	48.50	-3.01	5.03×10^3	47.96	-4.08
100	7.12×10^3	97.80	-2.20	6.28×10^3	97.81	-2.19	1.03×10^4	97.60	-2.40
150	1.10×10^4	152.02	1.35	9.76×10^3	151.92	1.28	1.60×10^4	152.20	1.47
<i>Correlation coefficient</i>	0.9996			0.9996			0.9995		
<i>Slope (counts s⁻¹ ($\mu\text{g dm}^{-3}$)⁻¹)</i>	7.25×10^1			6.43×10^1			1.06×10^2		
<i>Intercept (counts s⁻¹)</i>	2.06×10^1			-4.67×10^0			-3.45×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5028			0.2955			0.1827		
<i>Standard error</i>	2.1813			2.0399			2.3983		

Table 3.9: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when no internal standard was used.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	8.43×10^1	1.06	-	1.07×10^2	1.38	-	3.81×10^3	3.84	-
10	1.82×10^3	10.62	6.16	1.49×10^3	10.43	4.32	4.37×10^3	9.78	-2.21
50	8.63×10^3	48.18	-3.64	7.28×10^3	48.33	-3.35	7.83×10^3	46.76	-6.48
100	1.78×10^4	98.74	-1.26	1.49×10^4	98.00	-2.00	1.24×10^4	95.56	-4.44
150	2.74×10^4	151.41	0.94	2.31×10^4	151.86	1.24	1.79×10^4	154.05	2.70
<i>Correlation coefficient</i>	0.9997			0.9996			0.9981		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	1.81×10^2			1.53×10^2			9.37×10^1		
<i>Intercept (counts s^{-1})</i>	-1.08×10^2			-1.03×10^2			3.45×10^3		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.1836			0.2524			13.4835		
<i>Standard error</i>	1.6710			2.0298			4.5156		

Table 3.10: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when no internal standard was used.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	6.76×10^2	1.50	-	1.14×10^2	1.60	-	1.01×10^2	1.44	-
10	1.67×10^3	10.02	0.15	1.10×10^3	10.32	3.18	5.65×10^2	10.53	5.33
50	6.13×10^3	48.33	-3.35	5.46×10^3	48.67	-2.65	2.46×10^3	47.58	-4.84
100	1.20×10^4	98.79	-1.21	1.09×10^4	96.96	-3.04	5.08×10^3	99.02	-0.98
150	1.81×10^4	151.36	0.91	1.72×10^4	152.45	1.63	7.76×10^3	151.42	0.95
<i>Correlation coefficient</i>	0.9997			0.9994			0.9996		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	1.17×10^2			1.14×10^2			5.11×10^1		
<i>Intercept (counts s^{-1})</i>	5.01×10^2			-6.78×10^1			2.73×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	1.6930			0.3418			1.1743		
<i>Standard error</i>	1.6697			2.5588			1.9301		

Table 3.11: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when no internal standard was used.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	7.50×10^1	3.67	-	5.55×10^1	1.21	-	6.61×10^1	1.38	-
10	9.67×10^1	22.27	122.68	4.34×10^2	11.57	15.65	4.29×10^2	11.08	10.80
50	1.02×10^2	26.93	-46.14	1.74×10^3	47.20	-5.60	1.82×10^3	48.26	-3.47
100	1.88×10^2	100.77	0.77	3.58×10^3	97.58	-2.42	3.62×10^3	96.31	-3.69
150	2.53×10^2	156.36	4.24	5.59×10^3	152.44	1.63	5.73×10^3	152.96	1.98
<i>Correlation coefficient</i>	0.9776			0.9993			0.9991		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	1.17×10^0			3.66×10^1			3.74×10^1		
<i>Intercept (counts s^{-1})</i>	7.07×10^1			1.12×10^1			1.46×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	26.0360			1.6907			1.2346		
<i>Standard error</i>	15.3245			2.7999			3.0762		

Table 3.12: Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when no internal standard was used.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	6.44×10^1	0.47	-	6.87×10^1	0.32	-	1.24×10^2	1.40	-
10	3.62×10^2	11.36	13.61	1.47×10^2	10.15	1.51	4.68×10^3	10.85	8.53
50	1.40×10^3	49.34	-1.32	4.82×10^2	52.14	4.28	2.25×10^4	47.87	-4.26
100	2.68×10^3	96.08	-3.92	8.19×10^2	94.33	-5.67	4.65×10^4	97.67	-2.33
150	4.23×10^3	152.74	1.83	1.29×10^3	153.06	2.04	7.28×10^4	152.20	1.47
<i>Correlation coefficient</i>	0.9992			0.9985			0.9994		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	2.73×10^1			7.98×10^0			4.82×10^2		
<i>Intercept (counts s^{-1})</i>	5.14×10^1			6.61×10^1			-5.51×10^2		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	1.6948			4.9849			0.2284		
<i>Standard error</i>	2.9059			3.9211			2.4137		

Table 3.13: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when no internal standard was used.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.06×10^2	0.32	-	1.60×10^2	-0.39	-	8.42×10^1	0.73	-
10	4.29×10^2	11.23	12.29	2.88×10^2	12.26	22.63	7.95×10^2	11.03	10.28
50	1.49×10^3	47.14	-5.72	6.26×10^2	45.74	-8.52	3.36×10^3	48.20	-3.60
100	3.09×10^3	101.33	1.33	1.21×10^3	103.37	3.37	6.83×10^3	98.54	-1.46
150	4.53×10^3	149.98	-0.01	1.67×10^3	149.03	-0.65	1.05×10^4	151.51	1.00
<i>Correlation coefficient</i>	0.9996			0.9989			0.9997		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	2.96×10^1			1.01×10^1			6.90×10^1		
<i>Intercept (counts s^{-1})</i>	9.65×10^1			1.64×10^2			3.41×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.6533			7.1587			0.8686		
<i>Standard error</i>	1.9616			3.4455			1.7533		

Table 3.14: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when no internal standard was used.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.49×10^2	0.73	-	8.99×10^1	0.39	-	9.21×10^1	1.01	-
10	8.15×10^2	10.44	4.35	1.04×10^3	10.57	5.68	1.81×10^3	10.64	6.42
50	3.41×10^3	48.28	-3.44	4.69×10^3	49.48	-1.05	8.51×10^3	48.29	-3.42
100	6.96×10^3	100.04	0.04	9.26×10^3	98.27	-1.73	1.75×10^4	98.59	-1.41
150	1.04×10^4	150.52	0.35	1.42×10^4	151.29	0.86	2.69×10^4	151.47	0.98
<i>Correlation coefficient</i>	0.9999			0.9998			0.9997		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	6.86×10^1			9.37×10^1			1.78×10^2		
<i>Intercept (counts s^{-1})</i>	9.86×10^1			5.29×10^1			-8.83×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	1.3473			1.0966			0.2361		
<i>Standard error</i>	1.1463			1.3409			1.6850		

Table 3.15: Calibration data for ^{104}Ru when no internal standard was used.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.05×10^2	0.87	-
10	1.53×10^3	10.33	3.35
50	7.24×10^3	48.26	-3.48
100	1.50×10^4	99.92	-0.08
150	2.26×10^4	150.61	0.41
<i>Correlation coefficient</i>	0.9999		
<i>Slope (counts s^{-1} ($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	1.51×10^2		
<i>Intercept (counts s^{-1})</i>	-2.69×10^1		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5755		
<i>Standard error</i>	1.1944		

Table 3.16: Calibration data for Au and Ir when ^{36}Ar is used as internal standard.

	^{197}Au			^{191}Ir			^{193}Ir		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	5.64×10^{-6}	1.79	-	3.45×10^{-6}	1.66	-	3.61×10^{-6}	1.51	-
10	3.94×10^{-5}	11.02	10.23	3.36×10^{-5}	10.96	9.58	5.71×10^{-5}	11.56	15.64
50	1.76×10^{-4}	48.37	-3.25	1.56×10^{-4}	48.66	-2.69	2.52×10^{-4}	48.12	-3.76
100	3.46×10^{-4}	95.02	-4.98	3.06×10^{-4}	95.04	-4.96	5.00×10^{-4}	94.84	-5.16
150	5.61×10^{-4}	153.79	2.53	4.96×10^{-4}	153.69	2.46	8.15×10^{-4}	153.96	2.64
<i>Correlation coefficient</i>	0.9986			0.9986			0.9984		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	3.65×10^{-6}			3.24×10^{-6}			5.32×10^{-6}		
<i>Intercept</i>	-8.93×10^{-7}			-1.91×10^{-6}			-4.45×10^{-6}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.4840			0.2845			0.1758		
<i>Standard error</i>	3.9116			3.8119			4.1012		

Table 3.17: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when ^{36}Ar is used as internal standard.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	4.09×10^{-6}	1.54	-	5.20×10^{-6}	1.86	-	1.85×10^{-4}	3.85	-
10	9.02×10^{-5}	10.98	9.77	7.40×10^{-5}	10.80	7.97	2.17×10^{-4}	10.59	5.91
50	4.31×10^{-4}	48.34	-3.32	3.64×10^{-4}	48.48	-3.03	3.92×10^{-4}	47.38	-5.24
100	8.66×10^{-4}	95.96	-4.04	7.23×10^{-4}	95.23	-4.77	6.04×10^{-4}	92.04	-7.96
150	1.39×10^{-3}	153.18	2.12	1.17×10^{-3}	153.63	2.42	9.08×10^{-4}	156.14	4.09
<i>Correlation coefficient</i>	0.9990			0.9987			0.9961		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	9.13×10^{-6}			7.69×10^{-6}			4.75×10^{-6}		
<i>Intercept</i>	-1.00×10^{-5}			-9.07×10^{-6}			1.67×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.1768			0.2430	-		12.8949		
<i>Standard error</i>	3.2914			3.7481			6.3785		

Table 3.18: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when ^{36}Ar is used as internal standard.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.28×10^{-5}	1.92	-	5.53×10^{-6}	2.07	-	4.89×10^{-6}	1.90	-
10	8.28×10^{-5}	10.44	4.42	5.48×10^{-5}	10.69	6.86	2.81×10^{-5}	10.91	9.12
50	3.07×10^{-4}	48.55	-2.91	2.73×10^{-4}	48.83	-2.35	1.23×10^{-4}	47.76	-4.47
100	5.85×10^{-4}	95.90	-4.10	5.33×10^{-4}	94.22	-5.78	2.47×10^{-4}	96.21	-3.79
150	9.21×10^{-4}	153.19	2.12	8.76×10^{-4}	154.20	2.80	3.94×10^{-4}	153.21	2.14
<i>Correlation coefficient</i>	0.9990			0.9982			0.9989		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	5.87×10^{-6}			5.72×10^{-6}			2.57×10^{-6}		
<i>Intercept</i>	2.15×10^{-5}			-6.32×10^{-6}			-6.97×10^{-9}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.6288			0.3290			1.1305		
<i>Standard error</i>	3.3111			4.3593			3.3694		

Table 3.19: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when ^{36}Ar is used as internal standard.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.64×10^{-6}	3.47	-	2.69×10^{-6}	1.68	-	3.20×10^{-6}	1.84	-
10	4.80×10^{-6}	23.07	130.73	2.16×10^{-5}	11.92	19.25	2.13×10^{-5}	11.45	14.46
50	5.11×10^{-6}	28.22	-43.57	8.69×10^{-5}	47.37	-5.25	9.09×10^{-5}	48.42	-3.15
100	9.17×10^{-6}	96.55	-3.45	1.74×10^{-4}	94.81	-5.19	1.76×10^{-4}	93.57	-6.43
150	1.29×10^{-5}	158.69	5.79	2.84×10^{-4}	154.21	2.81	2.91×10^{-4}	154.72	3.14
<i>Correlation coefficient</i>	0.9773			0.9982			0.9978		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	5.95×10^{-8}			1.84×10^{-6}			1.88×10^{-6}		
<i>Intercept</i>	3.43×10^{-6}			-4.12×10^{-7}			-2.71×10^{-7}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	24.8009			1.6274			1.1882		
<i>Standard error</i>	15.4003			4.3943			4.8748		

Table 3.20: Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when ^{36}Ar is used as internal standard.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.12×10^{-6}	0.95	-	3.33×10^{-6}	0.73	-	6.03×10^{-6}	1.88	-
10	1.80×10^{-5}	11.73	17.28	7.31×10^{-6}	10.60	5.99	2.32×10^{-4}	11.21	12.07
50	7.00×10^{-5}	49.50	-0.99	2.41×10^{-5}	52.34	4.68	1.13×10^{-3}	48.02	-3.96
100	1.30×10^{-4}	93.31	-6.69	3.99×10^{-5}	91.46	-8.54	2.26×10^{-3}	94.93	-5.07
150	2.15×10^{-4}	154.51	3.01	6.54×10^{-5}	154.88	3.25	3.70×10^{-3}	153.96	2.64
<i>Correlation coefficient</i>	0.9978			0.9968			0.9984		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.38×10^{-6}			4.03×10^{-7}			2.43×10^{-5}		
<i>Intercept</i>	1.82×10^{-6}			3.04×10^{-6}			-3.97×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.6309			4.7914	-		0.2199		
<i>Standard error</i>	4.7953			5.8438			4.0889		

Table 3.21: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when ^{36}Ar is used as internal standard.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	5.13×10^{-6}	0.78	-	7.75×10^{-6}	-0.04	-	4.08×10^{-6}	1.21	-
10	2.13×10^{-5}	11.62	16.19	1.43×10^{-5}	12.77	27.73	3.95×10^{-5}	11.39	13.91
50	7.45×10^{-5}	47.36	-5.27	3.13×10^{-5}	46.14	-7.73	1.68×10^{-4}	48.37	-3.26
100	1.51×10^{-4}	98.40	-1.60	5.88×10^{-5}	100.09	0.09	3.32×10^{-4}	95.73	-4.27
150	2.30×10^{-4}	151.84	1.22	8.47×10^{-5}	151.04	0.69	5.32×10^{-4}	153.29	2.20
<i>Correlation coefficient</i>	0.9995			0.9993			0.9989		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.49×10^{-6}			5.10×10^{-7}			3.47×10^{-6}		
<i>Intercept</i>	3.97×10^{-6}			7.77×10^{-6}			-1.16×10^{-7}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.6289			6.8785			0.8362		
<i>Standard error</i>	2.3158			2.8093			3.4160		

Table 3.22: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when ^{36}Ar is used as internal standard.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	7.20×10^{-6}	1.20	-	4.36×10^{-6}	0.88	-	4.46×10^{-6}	1.50	-
10	4.04×10^{-5}	10.82	8.17	5.18×10^{-5}	10.93	9.33	8.97×10^{-5}	11.00	10.02
50	1.71×10^{-4}	48.47	-3.06	2.34×10^{-4}	49.63	-0.73	4.26×10^{-4}	48.45	-3.10
100	3.39×10^{-4}	97.18	-2.82	4.51×10^{-4}	95.48	-4.52	8.50×10^{-4}	95.81	-4.19
150	5.29×10^{-4}	152.33	1.56	7.22×10^{-4}	153.08	2.05	1.37×10^{-3}	153.25	2.16
<i>Correlation coefficient</i>	0.9994			0.9990			0.9989		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	3.46×10^{-6}			4.72×10^{-6}			8.97×10^{-6}		
<i>Intercept</i>	3.06×10^{-6}			1.99×10^{-7}			-8.96×10^{-6}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.2972			1.0557			0.2274		
<i>Standard error</i>	2.4372			3.2474			3.3519		

Table 3.23: Calibration data for ^{104}Ru when ^{36}Ar is used as internal standard.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	5.08×10^{-6}	1.36	-
10	7.59×10^{-5}	10.70	7.02
50	3.62×10^{-4}	48.43	-3.14
100	7.31×10^{-4}	97.10	-2.90
150	1.15×10^{-3}	152.41	1.61
<i>Correlation coefficient</i>	0.9994		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	7.58×10^{-6}		
<i>Intercept</i>	-5.21×10^{-6}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5542		
<i>Standard error</i>	2.5168		

Table 3.24: Calibration data for Au and Ir when ^{45}Sc is used as internal standard.

	^{197}Au			^{191}Ir			^{193}Ir		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.18×10^{-4}	0.58	-	7.25×10^{-5}	0.46	-	7.58×10^{-5}	0.33	-
10	8.79×10^{-4}	10.25	2.50	7.49×10^{-4}	10.17	1.73	1.27×10^{-3}	10.80	7.95
50	3.99×10^{-3}	49.91	-0.18	3.53×10^{-3}	50.19	0.39	5.71×10^{-3}	49.62	-0.76
100	7.75×10^{-3}	97.75	-2.25	6.85×10^{-3}	97.75	-2.25	1.12×10^{-2}	97.55	-2.45
150	1.20×10^{-2}	151.51	1.01	1.06×10^{-2}	151.42	0.95	1.74×10^{-2}	151.71	1.14
<i>Correlation coefficient</i>	0.9998			0.9998			0.9997		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	7.86×10^{-5}			6.96×10^{-5}			1.14×10^{-4}		
<i>Intercept</i>	7.32×10^{-5}			4.06×10^{-5}			3.82×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.4725			0.2777	-		0.1717		
<i>Standard error</i>	1.6081			1.5639			1.8085		

Table 3.25: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when ^{45}Sc is used as internal standard.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	8.58×10^{-5}	0.37	-	1.09×10^{-4}	0.69	-	3.88×10^{-3}	1.03	-
10	2.01×10^{-3}	10.18	1.84	1.65×10^{-3}	10.00	0.02	4.84×10^{-3}	10.41	4.10
50	9.80×10^{-3}	49.84	-0.32	8.26×10^{-3}	49.99	-0.01	8.90×10^{-3}	50.17	0.33
100	1.94×10^{-2}	98.69	-1.31	1.62×10^{-2}	97.95	-2.05	1.35×10^{-2}	95.44	-4.56
150	2.97×10^{-2}	150.91	0.61	2.50×10^{-2}	151.37	0.91	1.94×10^{-2}	152.96	1.97
<i>Correlation coefficient</i>	0.9999			0.9998			0.9990		
<i>Slope</i> $((\mu\text{g dm}^{-3})^{-1})$	1.96×10^{-4}			1.65×10^{-4}			1.02×10^{-4}		
<i>Intercept</i>	1.28×10^{-5}			-4.43×10^{-6}			3.77×10^{-3}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.1726			0.2372			12.5941		
<i>Standard error</i>	0.9573			1.4796			3.2028		

Table 3.26: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when ^{45}Sc is used as internal standard.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	6.88×10^{-4}	0.54	-	1.16×10^{-4}	0.89	-	1.03×10^{-4}	0.69	-
10	1.85×10^{-3}	9.71	-2.86	1.22×10^{-3}	9.89	-1.10	6.26×10^{-4}	10.15	1.46
50	6.96×10^{-3}	50.23	0.46	6.20×10^{-3}	50.36	0.73	2.79×10^{-3}	49.28	-1.45
100	1.31×10^{-2}	98.73	-1.27	1.19×10^{-2}	96.90	-3.10	5.54×10^{-3}	98.98	-1.02
150	1.97×10^{-2}	150.79	0.52	1.87×10^{-2}	151.95	1.30	8.41×10^{-3}	150.91	0.61
<i>Correlation coefficient</i>	0.9999			0.9995			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.26×10^{-4}			1.23×10^{-4}			5.53×10^{-5}		
<i>Intercept</i>	6.21×10^{-4}			6.16×10^{-6}			6.46×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.5897			0.3212			1.1033		
<i>Standard error</i>	0.9393			2.1875			0.9810		

Table 3.27: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when ^{45}Sc is used as internal standard.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	7.63×10^{-5}	-0.31	-	5.65×10^{-5}	0.47	-	6.73×10^{-5}	0.62	-
10	1.07×10^{-4}	23.80	138.01	4.81×10^{-4}	11.19	11.86	4.75×10^{-4}	10.69	6.88
50	1.16×10^{-4}	30.78	-38.43	1.97×10^{-3}	48.86	-2.27	2.07×10^{-3}	49.98	-0.05
100	2.05×10^{-4}	100.71	0.71	3.90×10^{-3}	97.53	-2.47	3.94×10^{-3}	96.25	-3.75
150	2.75×10^{-4}	155.01	3.34	6.06×10^{-3}	151.94	1.30	6.22×10^{-3}	152.46	1.64
<i>Correlation coefficient</i>	0.9821			0.9996			0.9993		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.28×10^{-6}			3.96×10^{-5}			4.05×10^{-5}		
<i>Intercept</i>	7.67×10^{-5}			3.78×10^{-5}			4.21×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	24.2390			1.5887			1.1601		
<i>Standard error</i>	13.7187			2.0643			2.6417		

Table 3.28: Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when ^{45}Sc is used as internal standard.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	6.55×10^{-5}	-0.35	-	7.00×10^{-5}	-0.87	-	1.27×10^{-4}	0.73	-
10	4.01×10^{-4}	10.98	9.78	1.63×10^{-4}	9.89	-1.14	5.18×10^{-3}	10.42	4.22
50	1.59×10^{-3}	51.15	2.30	5.48×10^{-4}	54.36	8.71	2.56×10^{-2}	49.50	-1.01
100	2.92×10^{-3}	96.00	-4.00	8.92×10^{-4}	94.20	-5.80	5.07×10^{-2}	97.63	-2.37
150	4.58×10^{-3}	152.22	1.48	1.40×10^{-3}	152.42	1.61	7.89×10^{-2}	151.72	1.15
<i>Correlation coefficient</i>	0.9993			0.9981			0.9997		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.96×10^{-5}			8.65×10^{-6}			5.22×10^{-4}		
<i>Intercept</i>	7.58×10^{-5}			7.74×10^{-5}			-2.57×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.5926			4.6804			0.2146		
<i>Standard error</i>	2.7855			4.4352			1.7865		

Table 3.29: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when ^{45}Sc is used as internal standard.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.08×10^{-4}	-0.55	-	1.63×10^{-4}	-1.94	-	8.57×10^{-5}	-0.01	-
10	4.75×10^{-4}	10.90	9.00	3.19×10^{-4}	12.29	22.86	8.80×10^{-4}	10.62	6.22
50	1.69×10^{-3}	48.92	-2.15	7.11×10^{-4}	48.06	-3.88	3.82×10^{-3}	49.91	-0.19
100	3.37×10^{-3}	101.29	1.29	1.32×10^{-3}	103.29	3.29	7.44×10^{-3}	98.48	-1.52
150	4.92×10^{-3}	149.44	-0.37	1.81×10^{-3}	148.30	-1.13	1.14×10^{-2}	151.00	0.67
<i>Correlation coefficient</i>	0.9999			0.9992			0.9999		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	3.21×10^{-5}			1.10×10^{-5}			7.47×10^{-5}		
<i>Intercept</i>	1.25×10^{-4}			1.84×10^{-4}			8.66×10^{-5}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.6137			6.7135			0.8162		
<i>Standard error</i>	1.1885			2.9688			1.1112		

Table 3.30: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when ^{45}Sc is used as internal standard.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.51×10^{-4}	-0.06	-	9.15×10^{-5}	-0.35	-	9.37×10^{-5}	0.32	-
10	9.02×10^{-4}	10.04	0.44	1.16×10^{-3}	10.14	1.35	2.00×10^{-3}	10.21	2.10
50	3.87×10^{-3}	50.03	0.07	5.32×10^{-3}	51.23	2.46	9.67×10^{-3}	49.96	-0.09
100	7.59×10^{-3}	99.99	-0.01	1.01×10^{-2}	98.21	-1.79	1.90×10^{-2}	98.54	-1.46
150	1.13×10^{-2}	149.99	0.00	1.54×10^{-2}	150.78	0.52	2.92×10^{-2}	150.98	0.65
<i>Correlation coefficient</i>	1.0000			0.9998			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	7.43×10^{-5}			1.01×10^{-4}			1.93×10^{-4}		
<i>Intercept</i>	1.56×10^{-4}			1.27×10^{-4}			3.19×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.2659			1.0306			0.2219		
<i>Standard error</i>	0.0482			1.3509			1.0402		

Table 3.31: Calibration data for ^{104}Ru when ^{45}Sc is used as internal standard.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.07×10^{-4}	0.17	-
10	1.69×10^{-3}	9.90	-0.96
50	8.22×10^{-3}	49.94	-0.12
100	1.64×10^{-2}	99.87	-0.13
150	2.46×10^{-2}	150.11	0.07
<i>Correlation coefficient</i>	1.0000		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	1.63×10^{-4}		
<i>Intercept</i>	7.85×10^{-5}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5408		
<i>Standard error</i>	0.1544		

Table 3.32: Calibration data for Au and Ir when ^{89}Y is used as internal standard.

	^{197}Au			^{191}Ir			^{193}Ir		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.67×10^{-4}	0.11	-	1.02×10^{-4}	0.00	-	1.07×10^{-4}	-0.13	-
10	1.23×10^{-3}	9.74	-2.57	1.05×10^{-3}	9.67	-3.31	1.78×10^{-3}	10.29	2.94
50	5.66×10^{-3}	49.92	-0.15	5.01×10^{-3}	50.21	0.41	8.10×10^{-3}	49.63	-0.74
100	1.12×10^{-2}	100.53	0.53	9.93×10^{-3}	100.52	0.52	1.62×10^{-2}	100.31	0.31
150	1.67×10^{-2}	149.69	-0.21	1.47×10^{-2}	149.60	-0.26	2.42×10^{-2}	149.90	-0.07
<i>Correlation coefficient</i>	1.0000			1.0000			1.0000		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.10×10^{-4}			9.78×10^{-5}			1.61×10^{-4}		
<i>Intercept</i>	1.54×10^{-4}			1.02×10^{-4}			1.28×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.4738			0.2785	...		0.1722		
<i>Standard error</i>	0.3907			0.4406			0.3392		

Table 3.33: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when ^{89}Y is used as internal standard.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.21×10^{-4}	-0.09	-	1.54×10^{-4}	0.23	-	5.46×10^{-3}	0.49	-
10	2.82×10^{-3}	9.68	-3.16	2.31×10^{-3}	9.50	-4.96	6.77×10^{-3}	9.58	-4.16
50	1.39×10^{-2}	49.84	-0.32	1.17×10^{-2}	50.00	0.01	1.26×10^{-2}	50.31	0.61
100	2.81×10^{-2}	101.47	1.47	2.35×10^{-2}	100.70	0.70	1.96×10^{-2}	99.08	-0.92
150	4.13×10^{-2}	149.10	-0.60	3.48×10^{-2}	149.56	-0.29	2.70×10^{-2}	150.54	0.36
<i>Correlation coefficient</i>	0.9999			1.0000			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.76×10^{-4}			2.32×10^{-4}			1.43×10^{-4}		
<i>Intercept</i>	1.45×10^{-4}			1.01×10^{-4}			5.39×10^{-3}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.1730			0.2379	-		12.6231		
<i>Standard error</i>	1.0164			0.5725			0.7369		

Table 3.34: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when ^{89}Y is used as internal standard.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	9.70×10^{-4}	0.06	-	1.63×10^{-4}	0.43	-	1.45×10^{-4}	0.22	-
10	2.59×10^{-3}	9.17	-8.30	1.71×10^{-3}	9.39	-6.07	8.76×10^{-4}	9.64	-3.64
50	9.87×10^{-3}	50.25	0.50	8.79×10^{-3}	50.38	0.76	3.96×10^{-3}	49.28	-1.45
100	1.90×10^{-2}	101.63	1.63	1.73×10^{-2}	99.64	-0.36	8.04×10^{-3}	101.79	1.79
150	2.74×10^{-2}	148.88	-0.74	2.60×10^{-2}	150.16	0.10	1.17×10^{-2}	149.07	-0.62
<i>Correlation coefficient</i>	0.9999			1.0000			0.9998		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.77×10^{-4}			1.73×10^{-4}			7.77×10^{-5}		
<i>Intercept</i>	9.58×10^{-4}			8.86×10^{-5}			1.27×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.5940			0.3221	-		1.1063		
<i>Standard error</i>	1.2468			0.5350			1.2623		

Table 3.35: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when ^{89}Y is used as internal standard.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.08×10^{-4}	-0.85	-	7.96×10^{-5}	0.02	-	9.48×10^{-5}	0.16	-
10	1.50×10^{-4}	22.75	127.46	6.73×10^{-4}	10.68	6.80	6.65×10^{-4}	10.18	1.83
50	1.65×10^{-4}	30.87	-38.25	2.80×10^{-3}	48.87	-2.26	2.93×10^{-3}	50.00	0.00
100	2.98×10^{-4}	105.10	5.10	5.66×10^{-3}	100.31	0.31	5.72×10^{-3}	99.00	-1.00
150	3.82×10^{-4}	152.13	1.42	8.43×10^{-3}	150.12	0.08	8.65×10^{-3}	150.65	0.44
<i>Correlation coefficient</i>	0.9828			0.9999			1.0000		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	1.80×10^{-6}			5.56×10^{-5}			5.69×10^{-5}		
<i>Intercept</i>	1.09×10^{-4}			7.88×10^{-5}			8.55×10^{-5}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	24.2669			1.5933			1.1637		
<i>Standard error</i>	13.4226			0.7854			0.7032		

Table 3.36: Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when ^{89}Y is used as internal standard.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	9.23×10^{-5}	-0.81	-	9.85×10^{-5}	-1.35	-	1.78×10^{-4}	0.28	-
10	5.61×10^{-4}	10.46	4.58	2.28×10^{-4}	9.31	-6.91	7.25×10^{-3}	9.93	-0.71
50	2.25×10^{-3}	51.19	2.38	7.76×10^{-4}	54.45	8.90	3.63×10^{-2}	49.50	-1.00
100	4.23×10^{-3}	98.78	-1.22	1.29×10^{-3}	97.10	-2.90	7.36×10^{-2}	100.37	0.37
150	6.38×10^{-3}	150.39	0.26	1.94×10^{-3}	150.50	0.33	1.10×10^{-1}	149.93	-0.05
<i>Correlation coefficient</i>	0.9999			0.9990			1.0000		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	4.16×10^{-5}			1.21×10^{-5}			7.33×10^{-4}		
<i>Intercept</i>	1.26×10^{-4}			1.15×10^{-4}			-2.49×10^{-5}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.5974			4.6946	.		0.2152		
<i>Standard error</i>	1.1437			3.2005			0.3961		

Table 3.37: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when ^{89}Y is used as internal standard.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.52×10^{-4}	-1.02	-	2.29×10^{-4}	-2.44	-	1.21×10^{-4}	-0.48	-
10	6.64×10^{-4}	10.36	3.62	4.46×10^{-4}	11.63	16.27	1.23×10^{-3}	10.11	1.12
50	2.40×10^{-3}	48.92	-2.16	1.01×10^{-3}	48.08	-3.84	5.41×10^{-3}	49.92	-0.17
100	4.89×10^{-3}	104.22	4.22	1.91×10^{-3}	106.59	6.59	1.08×10^{-2}	101.28	1.28
150	6.84×10^{-3}	147.53	-1.65	2.52×10^{-3}	146.14	-2.57	1.58×10^{-2}	149.17	-0.56
<i>Correlation coefficient</i>	0.9992			0.9978			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	4.50×10^{-5}			1.54×10^{-5}			1.05×10^{-4}		
<i>Intercept</i>	1.98×10^{-4}			2.67×10^{-4}			1.71×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.6152			6.7259			0.8185		
<i>Standard error</i>	2.9548			4.8410			0.9275		

Table 3.38: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when ^{89}Y is used as internal standard.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.13×10^{-4}	-0.53	=	1.29×10^{-4}	-0.82	-	1.32×10^{-4}	-0.14	-
10	1.26×10^{-3}	9.52	-4.76	1.62×10^{-3}	9.62	-3.76	2.80×10^{-3}	9.71	-2.90
50	5.49×10^{-3}	50.04	0.07	7.55×10^{-3}	51.25	2.50	1.37×10^{-2}	49.96	-0.08
100	1.10×10^{-2}	102.84	2.84	1.46×10^{-2}	101.00	1.00	2.76×10^{-2}	101.31	1.31
150	1.57×10^{-2}	148.13	-1.25	2.15×10^{-2}	148.94	-0.70	4.06×10^{-2}	149.16	-0.56
<i>Correlation coefficient</i>	0.9996			0.9999			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.04×10^{-4}			1.43×10^{-4}			2.71×10^{-4}		
<i>Intercept</i>	2.68×10^{-4}			2.45×10^{-4}			1.70×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.2692			1.0335			0.2225		
<i>Standard error</i>	2.0072			1.2224			0.9187		

Table 3.39: Calibration data for ^{104}Ru when ^{89}Y is used as internal standard.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.50×10^{-4}	-0.29	-
10	2.37×10^{-3}	9.40	-6.01
50	1.17×10^{-2}	49.94	-0.12
100	2.37×10^{-2}	102.68	2.68
150	3.42×10^{-2}	148.27	-1.15
<i>Correlation coefficient</i>	0.9997		
<i>Slope</i> $((\mu\text{g dm}^{-3})^{-1})$	2.29×10^{-4}		
<i>Intercept</i>	2.17×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.5423		
<i>Standard error</i>	1.8812		

Table 3.40: Calibration data for Au and Ir when ^{138}La is used as internal standard.

	^{197}Au			^{191}Ir			^{193}Ir		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.54×10^{-1}	3.01	-	1.56×10^{-1}	2.87	-	1.63×10^{-1}	2.73	-
10	1.76×10^0	11.98	19.80	1.50×10^0	11.92	19.16	2.56×10^0	12.50	25.01
50	7.65×10^0	46.98	-6.03	6.77×10^0	47.26	-5.48	1.09×10^1	46.75	-6.49
100	1.51×10^1	91.47	-8.53	1.34×10^1	91.50	-8.50	2.19×10^1	91.31	-8.69
150	2.61×10^1	156.56	4.37	2.30×10^1	156.45	4.30	3.79×10^1	156.71	4.47
<i>Correlation coefficient</i>	0.9957			0.9958			0.9955		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.68×10^{-1}			1.49×10^{-1}			2.45×10^{-1}		
<i>Intercept</i>	-2.51×10^{-1}			-2.72×10^{-1}			-5.05×10^{-1}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.4738			0.2785			0.1721		
<i>Standard error</i>	6.7496			6.6405			6.9172		

Table 3.41: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when ^{138}La is used as internal standard.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.84×10^{-1}	2.75	-	2.34×10^{-1}	3.06	-	8.33×10^0	5.58	-
10	4.04×10^0	11.93	19.28	3.31×10^0	11.75	17.53	9.71×10^0	11.87	18.66
50	1.88×10^1	46.98	-6.04	1.58×10^1	47.11	-5.78	1.70×10^1	45.18	-9.64
100	3.79×10^1	92.39	-7.61	3.16×10^1	91.70	-8.30	2.64×10^1	87.68	-12.32
150	6.46×10^1	155.95	3.97	5.45×10^1	156.38	4.25	4.22×10^1	159.69	6.46
<i>Correlation coefficient</i>	0.9964			0.9959			0.9906		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	4.20×10^{-1}			3.54×10^{-1}			2.20×10^{-1}		
<i>Intercept</i>	-9.72×10^{-1}			-8.48×10^{-1}			7.10×10^0		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.1731			0.2379			12.5505		
<i>Standard error</i>	6.1337			6.5650			9.9640		

Table 3.42: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when ^{138}La is used as internal standard.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.48×10^0	3.20	-	2.49×10^{-1}	3.27	-	2.20×10^{-1}	3.12	-
10	3.71×10^0	11.45	14.51	2.46×10^0	11.65	16.47	1.26×10^0	11.87	18.71
50	1.33×10^1	47.06	-5.88	1.19×10^1	47.43	-5.14	5.34×10^0	46.40	-7.19
100	2.56×10^1	92.22	-7.78	2.33×10^1	90.72	-9.28	1.08×10^1	92.60	-7.40
150	4.28×10^1	156.07	4.05	4.07×10^1	156.93	4.62	1.83×10^1	156.01	4.00
<i>Correlation coefficient</i>	0.9963			0.9952			0.9963		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.70×10^{-1}			2.63×10^{-1}			1.18×10^{-1}		
<i>Intercept</i>	6.12×10^{-1}			-6.12×10^{-1}			-1.49×10^{-1}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.5938			0.3221			1.1068		
<i>Standard error</i>	6.2606			7.1359			6.2202		

Table 3.43: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when ^{138}La is used as internal standard.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.64×10^{-1}	5.46	-	1.21×10^{-1}	2.90	-	1.44×10^{-1}	3.07	-
10	2.15×10^{-1}	23.99	139.93	9.65×10^{-1}	12.85	28.52	9.54×10^{-1}	12.39	23.93
50	2.22×10^{-1}	26.60	-46.81	3.78×10^0	46.03	-7.94	3.96×10^0	47.02	-5.96
100	4.01×10^{-1}	91.25	-8.75	7.61×10^0	91.25	-8.75	7.69×10^0	90.06	-9.94
150	5.98×10^{-1}	162.70	8.47	1.32×10^1	156.97	4.64	1.35×10^1	157.46	4.97
<i>Correlation coefficient</i>	0.9696			0.9951			0.9944		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.76×10^{-3}			8.48×10^{-2}			8.68×10^{-2}		
<i>Intercept</i>	1.49×10^{-1}			-1.25×10^{-1}			-1.22×10^{-1}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	24.0620			1.5929	.		1.1628		
<i>Standard error</i>	17.8035			7.2072			7.6676		

Table 3.44: Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when ^{138}La is used as internal standard.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.41×10^{-1}	2.22	-	1.50×10^{-1}	2.13	-	2.72×10^{-1}	3.07	-
10	8.05×10^{-1}	12.69	26.89	3.27×10^{-1}	11.66	16.63	1.04×10^1	12.15	21.45
50	3.05×10^0	48.01	-3.97	1.05×10^0	50.54	1.09	4.90×10^1	46.68	-6.64
100	5.69×10^0	89.79	-10.21	1.74×10^0	87.87	-12.13	9.90×10^1	91.41	-8.59
150	9.98×10^0	157.29	4.86	3.04×10^0	157.79	5.20	1.72×10^2	156.69	4.46
<i>Correlation coefficient</i>	0.9946			0.9933			0.9955		
<i>Slope</i> $((\mu\text{g dm}^{-3})^{-1})$	6.34×10^{-2}			1.86×10^{-2}			1.12×10^0		
<i>Intercept</i>	-1.70×10^{-4}			1.11×10^{-1}			-3.16×10^0		
<i>Detection limit</i> $(\mu\text{g dm}^{-3})$	1.5956			4.6824			0.2153		
<i>Standard error</i>	7.5661			8.4189			6.8868		

Table 3.45: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when ^{138}La is used as internal standard.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.31×10^{-1}	2.06	-	3.49×10^{-1}	1.45	-	1.84×10^{-1}	2.44	-
10	9.53×10^{-1}	12.58	25.82	6.40×10^{-1}	13.81	38.10	1.77×10^0	12.34	23.41
50	3.24×10^0	45.98	-8.05	1.36×10^0	44.52	-10.95	7.31×10^0	46.98	-6.03
100	6.58×10^0	94.64	-5.36	2.57×10^0	95.93	-4.07	1.45×10^1	92.15	-7.85
150	1.07×10^1	154.74	3.16	3.94×10^0	154.29	2.86	2.48×10^1	156.08	4.05
<i>Correlation coefficient</i>	0.9975			0.9974			0.9963		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	6.86×10^{-2}			2.35×10^{-2}			1.60×10^{-1}		
<i>Intercept</i>	9.01×10^{-2}			3.15×10^{-1}			-2.07×10^{-1}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.6157			6.7238			0.8186		
<i>Standard error</i>	5.0941			5.2013			6.2785		

Table 3.46: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when ^{138}La is used as internal standard.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.24×10^{-1}	2.44	-	1.96×10^{-1}	2.13	-	2.01×10^{-1}	2.71	-
10	1.81×10^0	11.79	17.92	2.32×10^0	11.90	19.04	4.02×10^0	11.95	19.54
50	7.42×10^0	47.06	-5.87	1.02×10^1	48.18	-3.64	1.85×10^1	47.08	-5.84
100	1.48×10^1	93.52	-6.48	1.97×10^1	91.91	-8.09	3.72×10^1	92.25	-7.75
150	2.46×10^1	155.18	3.45	3.36×10^1	155.87	3.91	6.35×10^1	156.01	4.01
<i>Correlation coefficient</i>	0.9973			0.9965			0.9964		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	1.59×10^{-1}			2.17×10^{-1}			4.13×10^{-1}		
<i>Intercept</i>	-6.42×10^{-2}			-2.67×10^{-1}			-9.17×10^{-1}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.2701			1.0335			0.2226		
<i>Standard error</i>	5.3568			6.0714			6.1948		

Table 3.47: Calibration data for ^{104}Ru when ^{138}La is used as internal standard.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.29×10^{-1}	2.58	-
10	3.40×10^0	11.67	16.66
50	1.57×10^1	47.06	-5.87
100	3.19×10^1	93.49	-6.51
150	5.35×10^1	155.21	3.47
<i>Correlation coefficient</i>	0.9973		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	3.49×10^{-1}		
<i>Intercept</i>	-6.69×10^{-1}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5428		
<i>Standard error</i>	5.3898		

Table 3.48: Calibration data for Au and Ir when ^{139}La is used as internal standard.

	^{197}Au			^{191}Ir			^{193}Ir		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.30×10^{-4}	0.47	-	2.02×10^{-4}	0.34	-	2.11×10^{-4}	0.21	-
10	2.36×10^{-3}	9.90	-1.03	2.01×10^{-3}	9.83	-1.71	3.42×10^{-3}	10.45	4.49
50	1.09×10^{-2}	49.65	-0.70	9.66×10^{-3}	49.93	-0.13	1.56×10^{-2}	49.37	-1.27
100	2.17×10^{-2}	99.59	-0.41	1.91×10^{-2}	99.58	-0.42	3.13×10^{-2}	99.37	-0.63
150	3.26×10^{-2}	150.40	0.27	2.88×10^{-2}	150.31	0.21	4.73×10^{-2}	150.60	0.40
<i>Correlation coefficient</i>	1.0000			1.0000			1.0000		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.15×10^{-4}			1.91×10^{-4}			3.13×10^{-4}		
<i>Intercept</i>	2.29×10^{-4}			1.36×10^{-4}			1.44×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.4799			0.2821			0.1744		
<i>Standard error</i>	0.4776			0.3745			0.6842		

Table 3.49: Calibration data for ^{102}Pd , ^{104}Pd and ^{105}Pd when ^{139}La is used as internal standard.

	^{102}Pd			^{104}Pd			^{105}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.39×10^{-4}	0.25	-	3.04×10^{-4}	0.56	-	1.08×10^{-2}	1.48	-
10	5.41×10^{-3}	9.85	-1.50	4.43×10^{-3}	9.67	-3.29	1.30×10^{-2}	9.32	-6.75
50	2.68×10^{-2}	49.58	-0.85	2.26×10^{-2}	49.73	-0.53	2.43×10^{-2}	49.78	-0.45
100	5.42×10^{-2}	100.52	0.52	4.53×10^{-2}	99.77	-0.23	3.78×10^{-2}	97.89	-2.11
150	8.07×10^{-2}	149.80	-0.13	6.81×10^{-2}	150.27	0.18	5.28×10^{-2}	151.53	1.02
<i>Correlation coefficient</i>	1.0000			1.0000			0.9997		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	5.38×10^{-4}			4.53×10^{-4}			2.80×10^{-4}		
<i>Intercept</i>	1.06×10^{-4}			4.87×10^{-5}			1.04×10^{-2}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.1753			0.2410			12.7873		
<i>Standard error</i>	0.4380			0.4547			1.7771		

Table 3.50: Calibration data for ^{106}Pd , ^{108}Pd and ^{110}Pd when ^{139}La is used as internal standard.

	^{106}Pd			^{108}Pd			^{110}Pd		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.92×10^{-3}	0.48	-	3.23×10^{-4}	0.77	-	2.86×10^{-4}	0.58	-
10	4.96×10^{-3}	9.29	-7.14	3.28×10^{-3}	9.56	-4.41	1.68×10^{-3}	9.79	-2.13
50	1.90×10^{-2}	49.95	-0.11	1.70×10^{-2}	50.11	0.21	7.63×10^{-3}	49.01	-1.98
100	3.66×10^{-2}	100.65	0.65	3.33×10^{-2}	98.71	-1.29	1.55×10^{-2}	100.84	0.84
150	5.35×10^{-2}	149.63	-0.25	5.09×10^{-2}	150.86	0.57	2.29×10^{-2}	149.79	-0.14
<i>Correlation coefficient</i>	1.0000			0.9999			0.9999		
<i>Slope (($\mu\text{g dm}^{-3}$)$^{-1}$)</i>	3.46×10^{-4}			3.37×10^{-4}			1.52×10^{-4}		
<i>Intercept</i>	1.75×10^{-3}			6.27×10^{-5}			1.98×10^{-4}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	1.6146			0.3263			1.1206		
<i>Standard error</i>	0.6607			1.0327			0.8390		

Table 3.51: Calibration data for ^{192}Pt , ^{194}Pt and ^{195}Pt when ^{139}La is used as internal standard.

	^{192}Pt			^{194}Pt			^{195}Pt		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.13×10^{-4}	0.58	-	1.57×10^{-4}	0.37	-	1.87×10^{-4}	0.52	-
10	2.88×10^{-4}	22.04	120.45	1.29×10^{-3}	10.82	8.21	1.28×10^{-3}	10.33	3.29
50	3.17×10^{-4}	30.44	-39.13	5.39×10^{-3}	48.61	-2.78	5.65×10^{-3}	49.72	-0.56
100	5.74×10^{-4}	103.66	3.66	1.09×10^{-2}	99.37	-0.63	1.10×10^{-2}	98.07	-1.93
150	7.48×10^{-4}	153.28	2.19	1.65×10^{-2}	150.83	0.55	1.69×10^{-2}	151.36	0.90
<i>Correlation coefficient</i>	0.9831			0.9999			0.9998		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	3.51×10^{-6}			1.09×10^{-4}			1.11×10^{-4}		
<i>Intercept</i>	2.10×10^{-4}			1.17×10^{-4}			1.29×10^{-4}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	24.5744			1.6137	-		1.1785		
<i>Standard error</i>	13.3378			1.1304			1.4151		

Table 3.52; Calibration data for ^{196}Pt , ^{198}Pt and ^{103}Rh when ^{139}La is used as internal standard.

	^{196}Pt			^{198}Pt			^{103}Rh		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	1.82×10^{-4}	-0.43	-	1.95×10^{-4}	-0.86	-	3.52×10^{-4}	0.61	-
10	1.08×10^{-3}	10.59	5.92	4.38×10^{-4}	9.39	-6.08	1.39×10^{-2}	10.09	0.95
50	4.34×10^{-3}	50.89	1.79	1.50×10^{-3}	54.07	8.15	6.99×10^{-2}	49.24	-1.52
100	8.15×10^{-3}	97.85	-2.15	2.49×10^{-3}	96.14	-3.86	1.42×10^{-1}	99.44	-0.56
150	1.25×10^{-2}	151.10	0.73	3.80×10^{-3}	151.26	0.84	2.15×10^{-1}	150.62	0.42
<i>Correlation coefficient</i>	0.9998			0.9989			0.9999		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	8.11×10^{-5}			2.37×10^{-5}			1.43×10^{-3}		
<i>Intercept</i>	2.17×10^{-4}			2.15×10^{-4}			-5.15×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.6177			4.7542	-		0.2180		
<i>Standard error</i>	1.5467			3.3730			0.7443		

Table 3.53: Calibration data for ^{96}Ru , ^{98}Ru and ^{99}Ru when ^{139}La is used as internal standard.

	^{96}Ru			^{98}Ru			^{99}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	3.00×10^{-4}	-0.62	-	4.53×10^{-4}	-1.81	-	2.39×10^{-4}	-0.12	-
10	1.28×10^{-3}	10.48	4.81	8.56×10^{-4}	11.59	15.93	2.36×10^{-3}	10.26	2.63
50	4.63×10^{-3}	48.65	-2.71	1.94×10^{-3}	47.74	-4.51	1.04×10^{-2}	49.64	-0.71
100	9.42×10^{-3}	103.22	3.22	3.68×10^{-3}	105.48	5.48	2.08×10^{-2}	100.33	0.33
150	1.34×10^{-2}	148.27	-1.15	4.93×10^{-3}	146.99	-2.01	3.09×10^{-2}	149.88	-0.08
<i>Correlation coefficient</i>	0.9995			0.9984			1.0000		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	8.79×10^{-5}			3.01×10^{-5}			2.05×10^{-4}		
<i>Intercept</i>	3.54×10^{-4}			5.07×10^{-4}			2.63×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	0.6231			6.8131			0.8290		
<i>Standard error</i>	2.2946			4.0773			0.3339		

Table 3.54: Calibration data for ^{100}Ru , ^{101}Ru and ^{102}Ru when ^{139}La is used as internal standard.

	^{100}Ru			^{101}Ru			^{102}Ru		
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	4.21×10^{-4}	-0.16	-	2.55×10^{-4}	-0.46	-	2.61×10^{-4}	0.20	-
10	2.42×10^{-3}	9.67	-3.28	3.10×10^{-3}	9.78	-2.16	5.38×10^{-3}	9.87	-1.26
50	1.06×10^{-2}	49.76	-0.48	1.46×10^{-2}	50.97	1.93	2.64×10^{-2}	49.69	-0.61
100	2.12×10^{-2}	101.87	1.87	2.82×10^{-2}	100.05	0.05	5.32×10^{-2}	100.37	0.37
150	3.08×10^{-2}	148.85	-0.76	4.20×10^{-2}	149.66	-0.23	7.94×10^{-2}	149.86	-0.09
<i>Correlation coefficient</i>	0.9998			1.0000			1.0000		
<i>Slope</i> ($(\mu\text{g dm}^{-3})^{-1}$)	2.04×10^{-4}			2.78×10^{-4}			5.29×10^{-4}		
<i>Intercept</i>	4.53×10^{-4}			3.82×10^{-4}			1.56×10^{-4}		
<i>Detection limit</i> ($\mu\text{g dm}^{-3}$)	1.2856			1.0467	-		0.2254		
<i>Standard error</i>	1.2925			0.6609			0.3174		

Table 3.55: Calibration data for ^{104}Ru when ^{139}La is used as internal standard.

^{104}Ru			
[Element] ($\mu\text{g dm}^{-3}$)	Intensity ratio	Calculated concentration ($\mu\text{g dm}^{-3}$)	% difference (certified value - calculated value)
0	2.97×10^{-4}	0.05	-
10	4.55×10^{-3}	9.56	-4.35
50	2.25×10^{-2}	49.67	-0.66
100	4.57×10^{-2}	101.73	1.73
150	6.68×10^{-2}	148.99	-0.67
<i>Correlation coefficient</i>	0.9999		
<i>Slope ($(\mu\text{g dm}^{-3})^{-1}$)</i>	4.47×10^{-4}		
<i>Intercept</i>	2.74×10^{-4}		
<i>Detection limit ($\mu\text{g dm}^{-3}$)</i>	0.5493		
<i>Standard error</i>	1.1972		