



5. CONCLUSIONS AND RECOMMENDATIONS

Chapter 5 presents conclusions drawn from the current research results and recommendations for clinical practice and for future research. A critical evaluation of the study is included,

5.1 INTRODUCTION

Objective measurement of middle-ear function continues to be refined and in recent years high frequency tympanometry has enjoyed added awareness and is increasingly being implemented in neonatal and infant hearing assessment protocols. In accordance with a dearth in current research, this study explored and validated the use of high frequency acoustic immittance measures in a large group of infants and described age- and gender related normative values.

To date, no large scale studies have been reported on high frequency tympanometry and acoustic reflex testing in the neonatal and infant population and limitations in subject size and the shortfall of age appropriate normative data in previous studies have been identified. Although the use of high frequency tympanometry has generally been proven and accepted as superior over conventional 226 Hz tympanometry in infants below 7 months of age (Lilly, 2005:24; Kei *et al.*, 2003:27; Margolis *et al.*, 2003:391; Purdy & Williams, 2000:10; Meyer *et al.*, 1997:194:); the need exists for classification systems and norms for interpreting these results to ensure effective diagnosis of middle ear pathology in infants (Sutton *et al.*, 1996:15).

Conclusions from the research findings in the current study will be presented in the following section to ascertain the significance of the obtained results and to determine implications for clinical practise.



5.2 CONCLUSIONS

This study has investigated and highlighted the use of high frequency probe tone (1000 Hz) tympanometry as a method of middle ear assessment in infants. Conclusions are discussed in Table 5.1 against the framework of the main and sub-aims formulated for the study.

TABLE 5.1 Conclusions according to sub-aims

• ***Sub aim 1 – Admittance (Y_a) tympanogram shape and characteristics within subgroups***

- ~ The total case sample of 936 ears was divided into two groups, depending on outcome of OAE testing. 869 ears (93%) that passed OAE-screening were considered to have normal middle ear functioning and were assigned to Group A. 67 ears (7%) failed OAE screening, were postulated to have possible middle ear pathology and were assigned to Group B.
- ~ 88% (n = 823 ears) of ears in the total case sample displayed *peaked* admittance tympanograms and in conjunction with an OAE pass, this was considered confirmation of normal middle ear functioning. *Flat* tympanograms, considered indicative of middle ear pathology, were recorded in 12% (n = 112) of the test ears.
- ~ Significantly high associations were observed between tympanogram shape and OAE outcome: 93% of ears with OAE-pass results (n = 869) displayed *peaked* 1000 Hz admittance tympanograms. In 7% of ears (60/869) in which flat tympanograms were recorded, OAE's were however still found to be present. 79% of ears failing OAE testing displayed *flat* (abnormal) tympanograms. Results of this study showed greater agreement between OAE failure and abnormal 1000 Hz tympanograms compared to results by Sutton *et al.*, (1996:12) who reported only about half of OAE fails (16/33) to correspond with abnormal 1000 Hz tympanograms.
- ~ This study has confirmed that abnormal 1000 Hz tympanometry is strongly associated with OAE failure, and that normal, peaked 1000 Hz tympanometry is strongly associated with OAE pass results.



- ~ 88% of Y_a tympanograms displayed discernable peaks. Double peaked tympanograms were recorded in 4.5% of the total sample and comprised 5% of ears (41) with peaked tympanograms ($n = 823$). A considerable *gender* effect was observed within the group displaying double peaked tympanograms, and 64% were male. The incidence of double peaked tympanograms were much higher in the present study compared with results by Kei *et al.*, (2003:24) who reported recording of double peaked tympanograms for only 3 out of 224 ears (1%). Differences in sample size and age of subjects could account for discrepancy in results.
 - ~ Statistically significant differences ($p = 0.00$) were observed in Y_a tympanograms between the group passing OAE testing (mean admittance 2.5 mmho) and the group failing OAE testing (mean admittance 1.78 mmho)
 - ~ For ears displaying a peaked Y_a tympanogram in conjunction with an OAE pass result, the mean value for uncompensated acoustic admittance (Y_a) tympanograms was recorded at 2.85 mmho, with a standard deviation of 1.13 mmho. Maximum admittance values corresponding to an OAE *pass* result were recorded at 9.64 mmho and minimum admittance relating to OAE *pass* result was 0.86 mmho.
 - ~ The mean pressure value in admittance tympanograms for ears *passing* OAE testing and displaying a peaked Y_a tympanogram, was 0.13 daPa with a standard deviation of 60.93 daPa. The highest extreme pressure value related to an OAE pass resulted was 185 daPa, and the lowest extreme was measured at -275 daPa. 87% of ears passing OAEs displayed peak pressure values bigger than -100 daPa and smaller and equal to 100daPa.
- **Sub aim 2 – Characteristics of susceptance (B_a) and conductance (G_a) tympanograms**
 - ~ Notching was mainly evident in the susceptance (B_a) tympanograms and this pattern was in agreement with Holte *et al.*, (1991:9) who found that reactance shifts towards mass with high probe frequencies and towards asymmetrical resistance, with the interactions resulting in notched susceptance (B_a) tympanograms.
 - ~ A great range of variation of B/G tympanogram shapes occurred across all age groups with the most frequent shape being the 3B1G and 1B1G Vanhuyse type.



- ~ Complex configuration and notching patterns were observed in susceptance and conductance tympanograms, making interpretation of B/G tympanograms more difficult (Holte *et al.*, 1991:22).
- ~ Mass and stiffness related elements can be assessed by 1000 Hz B/G tympanograms, but prior to more definite results and large scale investigations on the interpretation of B/G tympanograms, the use of **Y-admittance** tympanograms appear superior as a screening tool for middle ear functioning in infants to ease interpretation and classification. Highly significant associations between OAE pass results and single peaked admittance tympanograms (93%) and between OAE fail results and flat admittance tympanograms (79%) obtained in the current study suggest that admittance tympanograms have good sensitivity and specificity for assessment of middle ear status in infants. Kei *et al.*, (2003:27) reported single-peaked tympanograms, indicative of normal middle ear functioning, in 92.2 % of neonate ears.
- **Sub aim 3 – 1000 Hz probe tone acoustic reflexes**
 - ~ Ipsilateral acoustic reflexes at 1000 Hz were evaluated with a high frequency (1000 Hz) probe tone. Reflex thresholds were found to be present in 86% (n=760) of the ears tested.
 - ~ A mean threshold of 93 dB with a standard deviation of 9 dB and a 90% range of 80 – 105 dB was obtained.
 - ~ The higher percentage of present acoustic reflexes obtained in the current study compared to previous studies in infants was attributed to the fact that a 1000 Hz probe tone and an ipsilateral mid-frequency (1000 Hz) stimulus was used to activate the reflexes (Wetherby & Bennet 1980:107).
 - ~ Good agreement was observed between acoustic reflex presence or absence and normal and abnormal OAE and tympanometry results. This validated the use of OAE pass results and peaked admittance tympanograms as a control variable for normal middle ear functioning with identification of the subgroup of subjects used for the description of normative tympanometric values.
 - ~ Acoustic reflexes appear a good adjunct to 1000 Hz tympanometry for the confirmation of middle ear functioning in infants below seven months of age.



- **Sub aim 4 – High frequency immittance norms**

- ~ Age and gender specific normative values for 1000 Hz acoustic admittance, susceptance and conductance tympanograms were compiled for ears displaying peaked tympanograms in conjunction with OAE pass results.
- ~ An increase in mean admittance and standard deviation values were observed across age groups. Mean admittance for infants aged 0 – 4 weeks was 2.4 mmho, with a standard deviation of 0.8, compared to 2.9 mmho with standard deviation 1.0 and 3.8 mmho with standard of deviation of 1.5 for infants aged 5 – 28 and 29 – 52 weeks respectively.
- ~ In agreement to Palmu *et al.*, (2001:181) age specific normative values are needed for the interpretation of tympanometry obtained from infants and young children. Results obtained from this study may serve as a preliminary normative data for the interpretation of age specific tympanograms.

5.3 IMPLICATIONS OF FINDINGS

This study has provided normative values and results for 1000 Hz tympanometry and this may serve as a guide to further research and as preliminary norms for the identification of normal and abnormal high frequency tympanograms. However, as this study did not include results of abnormal tympanometry, more research on data of abnormal ears is needed to fully understand the effect of abnormal middle ear functioning on tympanograms recorded with a high frequency probe tone.

The prospect of employing high frequency tympanometry for timely identification and treatment of middle ear effusion in infants may serve as an important adjunct to tests currently used for audiological diagnosis in the infant population, to differentiate between middle ear pathology and sensori-neural hearing loss. This may also serve as a means to reduce high false positive test outcomes during neonatal hearing screening program relating to transient middle ear effusion. Thus high frequency tympanometry demonstrates promise for inclusion in neonatal audiological assessment procedures and hearing screening programs.



Previous studies investigating developmental changes and characteristics of high frequency tympanometry in infants have utilized susceptance and conductance tympanograms (Holte *et al.*, 1991:3, Meyer *et al.*, 1997:191) while others have utilized admittance tympanograms (Kei *et al.*, 2003:23, Margolis *et al.*, 2003:385, Meyer *et al.*, 1997:191, Sutton *et al.*, 1996:11, Thornton *et al.*, 1993:320) in analysis and description of results. Purdy and Williams (2000:19) reported that the majority of studies investigating high frequency tympanometry measured admittance (Y_a) or susceptance (B_a) tympanograms and that conductance (G_a) tympanograms had a limited diagnostic role in infants. Their recommended test protocol consequently suggested the use of susceptance or admittance tympanograms (Purdy *et al.*, 2000:22) for measurement of high frequency tympanograms in infants.

Conductance (G_a) tympanograms recorded in the present study showed notable similarities with admittance (Y_a) tympanograms when a simple visual analysis was applied (see Figure 4.12 for examples). Analysis of B/G tympanograms in terms of number of extrema can determine contribution of mass and stiffness related systems in the middle ear. The most frequent shapes observed in the current study were classified as 3B1G and 1B1G. The diagnostic value of the assessment of the mass- and stiffness related elements in the infant middle ear still remains to be determined (Purdy & Williams, 2000:22).

Owing to high associations observed between results of admittance tympanograms and OAE outcome, with a significant association of 93% between OAE pass results and peaked admittance tympanograms and of 79% between OAE fail results and flat admittance tympanograms, total admittance (Y_a) appears a valid method of categorisation of 1000 Hz tympanograms in infants. Due to greater variation in notching patterns that occur in susceptance and conductance tympanograms and complicates interpretation, an additional benefit of the use of admittance tympanograms is that it facilitates and simplifies interpretation as less variability in notching patterns occur.

5.4 CRITICAL EVALUATION OF RESEARCH PROJECT

A critical evaluation in the form of strengths and limitations of the current study is provided in Table 5.2.

TABLE 5.2 Strengths and limitations of the current study

STRENGTHS
<ul style="list-style-type: none"> ④ The study encompassed a large sample of subjects on which high frequency immittance measurements were performed, which allows for increased sensitivity of normative values that were compiled. ④ 1000 Hz probe tone tympanometry and acoustic reflexes were analyzed and age and gender specific norms were reported. ④ The use of an OAE pass result as a control variable for normal middle ear functioning proved successful and useful for identification of normal middle functioning and correlated well to peaked tympanogram results and present acoustic reflexes.
LIMITATIONS
<ul style="list-style-type: none"> ④ Poor infant co-operation prevented the whole test procedure to be performed on all infants. Breast or bottle feeding to pacify distressed infants in addition to visual distraction proved successful in enhancing infant co-operation, though 100% success rate could not be obtained. ④ Due to failure of subjects to return for follow-up OAE screening it had not been possible to confirm or reject sensori-neural hearing loss in cases with a combination of a peaked tympanogram and absent OAEs. A further limitation was due to the fact that ABR screening results were not considered in the analysis of the results for subjects that underwent AABR testing, ④ Uncompensated acoustic admittance and tympanometric peak pressure were the only variables analysed in this study. For more detailed analysis and for comparison between studies tympanometric width, compensated static admittance and tympanometric gradient could be included as measurement variables in subsequent investigations.



5.5 CLINICAL GUIDELINES FOR INTERPRETATION OF HIGH FREQUENCY IMMITTANCE MEASURES

- ~ High frequency tympanometry in combination with acoustic reflexes proves a useful and sensitive measure of middle ear functioning in infants. *Peaked tympanograms and present acoustic reflexes strongly indicate normal middle ear functioning.*
- ~ High frequency admittance (Y_a) tympanograms are easier to interpret and proves a suitable measure if the assessment of mass and stiffness-related elements are not the primary objective.
- ~ In agreement with previous reports, double peaked tympanograms indicate good agreement with OAE pass results, and can therefore be interpreted as normal.
- ~ OAE pass results were highly associated with tympanic peak pressures values *greater than -100 daPa and smaller than 100 daPa*, with 87% of ears in the OAE pass group falling within this region. Positive middle ear pressure >150 daPa may be an important indicator of the presence of middle ear effusion in infants.
- ~ Age related normative data described in this study offer guidelines for the interpretation of high frequency tympanometry and reflexes in infants.

5.6 RECOMMENDATIONS FOR FUTURE RESEARCH

- Further research is needed on larger subgroups with abnormal middle ear functioning for comparisons to be made between results of normal and abnormal measures of high frequency tympanometry.
- More research on the validation of admittance tympanograms in relation to OAE outcome and other tests of middle ear assessment is necessary to further validate the use high frequency (1000 Hz) Y_a tympanograms as a single indicator of middle ear functioning in infants.



- Classification systems based on single component admittance tympanograms should be explored and developed to ease classification and interpretation of 1000 Hz tympanograms. This could provide a more scientific base for interpretation of high frequency tympanometry opposed to the current practise employing a simple pass criteria based on the presence of any peaked tympanometric pattern.

5.7 FINAL COMMENTS

The prospect of employing high frequency tympanometry to assess middle ear functioning in infants, addresses one of the major challenges of early identification and differentiation between middle ear pathology and sensori-neural hearing loss in infants. High frequency admittance tympanograms prove useful as part of a paediatric audiologic test battery for infants less than twelve months of age, though it is important to remember that concerns about hearing sensitivity after referral from newborn hearing screen cannot be dismissed on the basis of a flat tympanogram (Holte & Margolis, 2002:390). 1000 Hz admittance tympanometry proves a valid and useful method for identification middle ear pathology in infants. Normative values, as described in this study, offer guidelines for further research into universal normative values and a classification system for high frequency acoustic immittance measures in infants.

5.8 SUMMARY

Chapter 5 provided conclusions and recommendations based on the results obtained in this study. Significant findings were highlighted and recommendations for clinical practise provided. High frequency tympanometry shows great promise for timely diagnosis of middle ear dysfunction in infants and for differentiation between true and false positive results from hearing screening.



REFERENCE LIST

1. Babbie, E. & Mouton, J. 1998. *The Practice of Social Research*. Oxford University Press Southern Africa, Cape Town
2. Bhoola, D. & Hugo, R. 1995, 'Prevalence: outer and middle ear disorders in black and Indian preschool children from Durban', *South African Journal of Communication Disorders*, vol. 42, pp. 19 – 27
3. Boone, R.T., Bower, C.M. & Martin, P.F. 2005, 'Failed newborn Hearing Screens as presentation for Otitis Media with Effusion in the Newborn population', *International Journal of Paediatric Otorhinolaryngology*, vol. 69, pp. 393 – 397
4. Casselbrant, M.L., Gravel, J.S., Margolis, R.H. & Marchisio, P. 2002, 'Diagnosis and Screening', *The Annals of Otology, Rhinology & Laryngology*, vol. 111, no. 3, pp. 95 – 101
5. Davies, A., Bramford, J. & Stevens, J. 2001, 'Performance of neonatal and infant hearing screens: sensitivity and specificity', *British Journal of Audiology*, vol. 35, no. 1, pp. 3 – 15
6. De Chicchis, A.R., Wendell Todd, N. & Nozza, R.J. 2000, 'Developmental Changes in Aural Acoustic Admittance Measurements', *Journal of the American Academy of Audiology*, vol. 11, no. 2, pp. 97 – 102
7. DeConde Johnson, C. 2002, In Katz, J. (ed.) *Handbook of Clinical Audiology*, 5th ed., Lippincot Williams & Wilkins, Baltimore, Chapter 24
8. Doyle, K., Burggraaff, B., Fujikawa, S., Kim, J., & MacArthur, C. 1997, 'Neonatal hearing screening with otoscopy, auditory brainstem response and oto-acoustic emissions', *Otolaryngology-Head and Neck Surgery*, vol. 116, pp. 597-603

9. Durrheim, K. 1999, *Research in Practice: Applied Methods for the Social Sciences*, University of Cape Town Press, South Africa, Chapter 3 & 6
10. El-Refaie, A., Parker, D.J. & Bamford, J.M. 1996, 'Otoacoustic emission versus ABR screening: The effect of external and middle ear abnormalities in a group of SCBU neonates', *British Journal of Audiology*, vol. 30, pp. 3 – 8
11. Engel, J., Mahler, E., Anteunis, L., Marres, E. & Zielhuis, G. 2001, 'Why are NICU infants at risk for chronic otitis media with effusion?', *International Journal of Pediatric Otorhinolaryngology*, vol. 57, pp. 137 – 144
12. Feeney, M.P. 2005, 'Wideband Energy Reflectance', *ASHA Leader*, vol. 10, no. 5, pp. 6-7, 24-25
13. Feeney, M.P. & Grant, I.L. 2003, 'Wideband Energy Reflectance Measurements in Adults with Middle-Ear Disorders', *Journal of Speech, Language, and Hearing Research*, vol. 46, pp. 901 – 911
14. Feeney, M.P., Keefe, D.H. & Sanford, C.A. 2004, 'Wideband Reflectance Measures of the Ipsilateral Acoustic Stapedius Reflex Threshold', *Ear and Hearing*, vol. 25, pp. 421 – 430
15. Ferekidis, E. 2003, 'The Use of Advanced Tympanometry Techniques in the Differential Diagnosis of Middle Ear Pathology', *ENT News*, vol. 11, no. 6, pp. 59 - 60
16. Fowler, C.G. & Shanks, J.E. 2002, 'Tympanometry', In Katz, J. (ed.) *Handbook of Clinical Audiology*, 5th ed., Lippincot Williams & Wilkins, Baltimore, Chapter 12



17. Gaihede, M. & Ovesen, T. 1997, 'Precision of Tympanometric Measurements', *Journal of Speech, Language and Hearing Research*, vol. 40, pp. 215 – 222
18. Gelfand, S.A. 2002, 'The Acoustic Reflex', In Katz, J. (ed.) *Handbook of Clinical Audiology*, 5th ed., Lippincot Williams & Wilkins, Baltimore, Chapter 13
19. Gliddon, M.L. & Sutton, G.J. 2001, 'Prediction of 8-month MEE from neonatal risk factors and test results in SCBU and full-term babies', *British Journal of Audiology*, vol. 35, no. 1, pp. 77 – 85
20. Goddard, W. & Melville, S. 2005, *Research Methodology: An Introduction*, 2nd edition, Juta & Co. Ltd, Lansdowne
21. Govender, R. 1998, *Pneumatic otoscopy and tympanometry in the identification of middle ear effusion*, Masters Thesis, University of Pretoria, South Africa
22. Gruber, J. 2002, 'Basic Multifrequency Tympanometry: The Physical Background' pp. 1 - 21 available online at <http://www.Lymenet.de/symptoms/tympanom/basic tym.htm>
23. Grason-Stadler, 2000, *GSI@TymStar Version 2 Middle-Ear Analyzer Reference Instruction Manual*, Madison
24. Hall III, J.W. & Mueller III, H.G. 1997, *Audiologists' Desk Reference Vol. 1*, Singular Publishing Group, Inc. London
25. Hirsch, J.E., Margolis, R.H. & Rykken, J.R. 1992, 'A Comparison of Acoustic Reflex and Auditory Brain Stem Response Screening of High-Risk Infants', *Ear and Hearing*, vol. 13, no. 3, pp. 181 – 186

26. Hogan, S.C., Stratford, K.J. & Moore, D.R. 1997, 'Duration and recurrence of otitis media with effusion in children from birth to 3 years: prospective study using monthly otoscopy and tympanometry', *British Medical Journal*, vol. 314, no. 7077, pp. 350 – 356
27. Holte, L., & Margolis, R.H. 2002, 'Contemporary research in tympanometry', *Current Opinion in Otolaryngology & Head and Neck Surgery*, vol. 10, no. 5, pp. 387 – 391
28. Holte, L., Margolis, R.H. & Cavanaugh, Jr. R.M. 1991, 'Developmental Changes in Multifrequency Tympanograms', *Audiology*, vol. 30, pp. 1 – 24
29. Jerger, J. & Northern, J.L. 1980, *Clinical Impedance Audiometry*, 2nd Edition, American Electromedics Corporation
30. Joint Committee on Infant Hearing 2000, *Year 2000 position statement: Principles and Guidelines for Early Hearing Detection and Intervention Program*, *Audiology Today*, Special Issue August 2000, pp. 6 - 27
31. Joint Committee on Infant Hearing, *1994 Position Statement*, available online at <http://www.jcih.org>, pp. 1 - 4
32. Keefe, D.H. & Levi, E. 1996, 'Maturation of the Middle and External Ears: Acoustic Power-Based Responses and Reflectance Tympanometry', *Ear & Hearing*, vol. 17, no. 5, pp. 361 – 373
33. Keefe, D.H., Folsom, R.C., Gorga, M.P., Vohr, B.R., Bulen, J.C. & Norton, S.J. 2000, 'Identification of Neonatal Hearing Impairment: Ear-Canal Measurements of Acoustic Admittance and Reflectance in Neonates', *Ear & Hearing*, vol. 21, no. 5, pp. 443 – 461



34. Keefe, D.H., Zhao, F., Neely, S.T., Gorga, M.P. & Vohr, B.R. 2003, 'Ear-canal acoustic admittance and reflectance effects in human neonates. I. Predictions of otoacoustic emission and auditory brainstem responses', *Journal of the Acoustic Society of America*, vol. 113, no. 1, pp 389 – 406
35. Kei, J., Allison-Levick, J., Dockray, J., Harrys, R., Kirkegard, C., Wong, J., Maurer, M., Hegarty, J., Young, J. & Tudehope, D. 2003, 'High-Frequency (1000Hz) Tympanometry in Normal Neonates', *Journal of the American Academy of Audiology*, vol. 14, no. 1, pp. 20 – 28
36. Koivunen, P., Uhari, M., Laitakari, K., Alho, O-P. & Luotonen, J. 2000, 'Otoacoustic Emissions and Tympanometry in Children with Otitis Media', *Ear & Hearing*, vol. 21, no. 3, pp. 212 – 217
37. Lantz, J., Petrak, M. & Prigge, L. 2004, 'Using the 1000 Hz probe tone for immittance measurements in infants', *The Hearing Journal*, vol. 57, no. 10, pp. 34 – 42
38. La Rossa, M. M., Mitchell, S. & Warnecke Cardinal, J. 1993, 'Tympanometry as a Screening Tool in the NICU: Is it Effective?' *Neonatal Network*, vol. 12, no. 8, pp. 32 – 34
39. Leedy, P.D. 1993, *Practical Research: Planning and Design*, 5th ed., Macmillan Publishing Company, USA
40. Leedy, P.D. & Ormrod, J.E. 2001, *Practical Research: Planning and Design*, 7th ed., Prentice-Hall, Inc. USA
41. Lilly, D. 2005, 'The Evolution of Aural Acoustic-Immittance Measurements', *ASHA Leader*, vol. 10, no. 5, pp. 6, 24
42. Luterman, D.M., Kurtzer-White, E. & Seewald, R.C. 1999, *The Young Deaf Child*, York Press, Inc. Baltimore, Maryland



43. Marchant, C.D., McMillan, P.M., Shurin, P.A., Johnson C.E., Turczyk, V.A., Feinstein, J.C. & Panek, D.M. 1984, 'Objective diagnosis of otitis media in early infancy by tympanometry and ipsilateral acoustic reflex thresholds', *Journal of Pediatrics*, vol. 109, pp. 590 – 595
44. Margolis, R.H., Bass-Ringdahl, S., Hanks, W.D., Holte, L. & Zapala, D.A. 2003 'Tympanometry in Newborn Infants – 1 kHz Norms', *Journal of the American Academy of Audiology*, vol. 14, no. 7, pp. 383 – 392
45. Marieb, E.N. 1995, *Human Anatomy and Physiology*, 3rd Edition, The Benjamin-Cummings Publishing Company Inc., California
46. Martin, F.N. & Clark, J.G. 2000, *Introduction to Audiology*, 7th Edition, Allyn & Bacon, Pearson Education, USA
47. Mauk, G.W. & White, K.R. 1995, 'Giving Children a Sound Beginning: The Promise of Universal Newborn Hearing Screening', *The Volta Review*, vol. 97, pp. 5 – 32
48. McMillan, P.M., Marchant, C.D. & Shurin, P.A. 1985, 'Ipsilateral acoustic reflexes in infants', *Annals of Otolaryngology, Rhinology & Laryngology*, vol. 94, no.2, pp. 145 – 148
49. Mencher, G.T., Davis, A. C., DeVoe, S.J. & Beresford, D. 2001, 'Universal Neonatal Hearing Screening: Past, Present, and Future', *American Journal of Audiology*, vol. 10, pp. 3 – 12
50. Meyer, S. E., Jardine, C.A. & Deverson, W. 1997, 'Developmental changes in tympanometry: A case study', *British Journal of Audiology*, vol. 31, pp. 189 – 195



51. MRC Multi-Centre Otitis Media Study Group. 1999, 'Sensitivity, Specificity and predictive value of tympanometry in predicting a hearing impairment in otitis media effusion', *Clinical Otolaryngology*, vol. 24, pp. 294 – 3000
52. Neuman, W.L., 1997, *Social Research Methods: Qualitative and Quantitative Approaches*, 3rd edition, Boston: Allyn and Bacon
53. Neumann, J., Uppenkamp, S. & Kollmeier, B. 1996, 'Detection of the acoustic reflex below 80 dB HL', *Audiology and Neurootology*, vol. 1, pp. 359 – 369
54. Northern, J.L. & Downs, M.P. 1984, *Hearing in Children*, 3rd ed. Baltimore, Lippincot Williams & Wilkens
55. Northern, J.L. & Downs, M.P. 2002, *Hearing in Children*, 5th ed. Baltimore, Lippincot Williams & Wilkens
56. Nozza, R.J., Bluestone, C.D., Kardatzke, D. & Bachman, R. 1994,' Identification of Middle Ear Effusion by Aural Acoustic Admittance and Otoscopy', *Ear and Hearing*, vol. 15, issue 4, pp. 310 – 323
57. Palmu, A., Puhakka, H., Huhtala, H., Takala, A.K. & Kilpi, T. 2001, 'Normative Values for Tympanometry in 7- and 24-month old Children', *Audiology*, vol. 40, pp. 178 – 184
58. Palmu, A., Puhakka, H., Rahko, T. & Takala, A. 1999, 'Diagnostic value of tympanometry in infants in clinical practice', *Internation Journal of Pediatric Otorhinolaryngology*, vol. 49, pp. 207 – 213
59. Palmu, A., Puhakka, H., Rahko, T., Takala, A. & Kilpi, T. 2002, 'Predicting the development and outcome of otitis media by tympanometry', *International Journal of Pediatric Otorhinolaryngology*, vol. 62, pp. 135 – 142



60. Palmu, A., Rahko, T., Puhakka, H. & Takala, A.K. 2000, 'Interrator Agreement on tympanometry in infants', *Scandanavian Audiology*, vol. 29, pp. 260 – 265
61. Pellet, F.S., Cox, L.C. & MacDonald, C.B. 1997, 'Use of Acoustic Reflectometry in the Detection of Middle Ear Effusion', *Journal of the American Academy of Audiology*, vol. 8, pp. 181 – 187
62. Petrak, M. 2002. *Tympanometry beyond 226 Hz – What's different in Babies?*, ICS Medical, Illinois.
63. Purdy, S.C. & Williams, M. J. 2000, 'High frequency tympanometry: A valid and reliable immittance test protocol for young infants?', *The New Zealand Audiological Society Bulletin*, vol. 10, no. 3, pp. 9 – 24
64. Reynolds, L. & Morton, L.P. 1995, 'Acoustic reflex measures at high probe frequency: a phasor diagram approach', *British Journal of Audiology*, vol. 29, no. 3, pp. 144 – 152
65. Rhodes, M.C., Margolis, R.H., Hirsch, J.E. & Napp, A.P. 1999, 'Hearing Screening in the Newborn Intensive Care Nursery: Comparison of Methods', *Otolaryngology, Head and Neck Surgery*, vol. 120, pp. 799 – 808
66. Rossetti, L.M. 1996, *Communication Intervention: Birth to Three*, Singular Publishing Group Inc, San Diego
67. Roush, J. 2001, *Screening for Hearing Loss and Otitis Media In Children*, Singular Thomson Learning, Canada.
68. Roush, J., Drake, A. & Sexton, J.E. 1992, 'Identification of Middle Ear Dysfunction in Young Children: A Comparison of Tympanometric Screening Procedures', *Ear and Hearing*, vol. 13, no. 2, pp. 63 - 69



69. Shahnaz, N. 2002, *Immittance Principles*, University of British Columbia, School of Audiology and Speech Sciences, available online at <http://www.audiospeech.ubc.ca/school/faculty/navid/Imported%20Files/MFT.doc>
70. Shahnaz, N. 2003, *Immittance in Newborns*, University of British Columbia, School of Audiology and Speech Sciences, available online at <http://www.audiospeech.ubc.ca/school/faculty/navid/Imported%20Files/Newborn.doc>
71. Shapiro, N.L. & Novelli, V. 1998, 'Otitis media in children with vertically-acquired HIV infection: the Great Ormond Street Hospital experience', *International Journal of Pediatric Otorhinolaryngology*, vol. 45, pp. 69 – 75
72. Silman, S. & Silverman, C.A. 1991, *Auditory Diagnosis: Principles and Applications*, Academic Press, Inc., San Diego, California
73. Sininger, Y.S. 2003, 'Audiologic assessment in infants', *Current Opinion in Otolaryngology & Head and Neck Surgery*, vol. 11, pp. 378 – 382
74. Sutton, G.J. (ed.) 2000, 'Newborn Hearing Screening Programme. Tympanometry in neonates and infants under 4 months: A recommended Test Protocol', available online at <http://www.unhs.org.uk/documents/workbook/Chapter3/03TympanometryProtocol>
75. Sutton, G.J., Gleadle, P. & Rowe, S.J. 1996, 'Tympanometry and otoacoustic emissions in a cohort of special care neonates', *British Journal of Audiology*, vol. 30, pp. 9 – 17
76. Swanepoel, D.C.D, 2004, *Infant hearing screening at maternal and child health clinics in a developing South African community*, Doctoral Thesis, University of Pretoria, South Africa
77. Terre Blanche, M. & Durrheim, K. 1999, *Research in Practice: Applied Methods for the Social Sciences*, University of Cape Town Press, South Africa



78. Thornton, A.R.D., Kimm, L., Kennedy, C.R. and Cafarelli-Dees, D. 1993, 'External- and middle ear factors affecting evoked acoustic emissions in neonates', *British Journal of Audiology*, vol. 27, pp. 319 – 327
79. Trine, M.B., Hirsch, J.E. & Margolis, R.H. 1993, 'The Effect of Middle Ear Pressure on Transient Evoked Otoacoustic Emissions', *Ear and Hearing*, vol. 14, no. 6, pp. 401 – 407
80. Wada, H., Ohyama, K., Kobayashi, T., Koike, T. & Noguchi, S. 1995, 'Effect of Middle Ear on Otoacoustic Emissions', *Audiology*, vol. 34, pp. 161 – 176
81. Weatherby, L.A. & Bennett, M.J. 1980, 'The Neonatal Acoustic Reflex', *Scandinavian Audiology*, vol. 9, pp. 103 – 110
82. Wiley, T.L. & Fowler, C.G. 1997, *Acoustic Immittance Measures in Clinical Audiology: A Primer*, Singular Publishing Group, Inc., California
83. Wiley, T.L. & Stoppenbach, D.T. 2002, 'Basic Principles of Acoustic Immittance Measures', In Katz, J. (ed.) *Handbook of Clinical Audiology*, 5th ed., Lippincot Williams & Wilkins, Baltimore, Chapter 11
84. Wisker, G. 2001. *The Postgraduate Research Handbook*. Palgrave, Hampshire.
85. Woods, C.R. 2003, 'Lack of Association of Race / Ethnicity and Otitis Median in the First 2 Years of Life', *Clinical Pediatrics*, vol. 42, no. 8, pp. 687 – 696
86. Yeo, S.W, Park, S.N., Park Y.S. & Suh, B.D. 2002, 'Effect of middle ear effusion on otoacoustic emissions', *The Journal of Laryngology & Otology*, vol. 116, pp. 794 – 799



87. Yoshinaga-Itano, C. 2001. Universal Newborn Hearing Screening: Speech, Language and Social-Emotional Outcomes for Children with Congenital Hearing Loss, Presentation at the Phonak International Paediatric Conference, South Africa



APPENDIX A: Data recording sheet

DATA SHEET

ID NO.																			
--------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Respondent no					V1					1-4
---------------	--	--	--	--	----	--	--	--	--	-----

SECTION A ~ IDENTIFYING INFORMATION

a) Gender

Male	1	Female	2
------	---	--------	---

V2		5
----	--	---

b) Child's Age

Weeks		
-------	--	--

V3			6-7
----	--	--	-----

c) Mother's Age

Years		
-------	--	--

V4			8-9
----	--	--	-----

d) Home Language

Tswana	1
Sepedi	2
Shangaan	3
Zulu	4
English	5
Afrikaans	6
Other	7

V5		10
----	--	----

e) Race

Black	1
Coloured	2
Indian	3
White	4

V6		11
----	--	----

f) Primary Caregiver

Mother	1
Father	2
Both	3
Grandparents	4
Extended family	5
Foster parents	6

V7		12
----	--	----



g) Educational Qualifications

i. Biological Mother

< St. 6	1
St. 6-8	2
St. 8-10	3
Diploma/Degree	4
Postgraduate	5

V8 13

ii. Biological Father

< St. 6	1
St. 6-8	2
St. 8-10	3
Diploma/Degree	4
Postgraduate	5

V9 14

h) Average Household Income (p/m)

<R500	1
R501 – R1000	2
R1001 – R2000	3
R2001 – R 5000	4
R5000+	5

V10 15

i) No. of children (Biological mother)

Born	<input type="text"/>	<input type="text"/>
Still living	<input type="text"/>	<input type="text"/>

V11 16-17
V12 18-19

j) Marital status of Biological parents

Married	1
Never married	2
Divorced	3
Widow	4

V13 20

k) Housing

Own house/flat	1
Informal housing	2
Renting	3
With others	4

V14 21

SECTION B ~ RISK INDICATORS

a) Family History of childhood Hearing loss

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V15 22



b) Hyperbilirubinemia

Levels requiring blood transfusion/exchange

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V16 23

If levels are known, are they in excess of the following amounts,

Birth weight (grams)	Bili level
≤ 1000	10.0
1001 – 1250	10.0
1251 – 1500	13.0
1501 – 2000	15.0
2001 – 2500	17.0
2500 +	18.0

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V17 24

c) Congenital infections

Yes	1	No	2
-----	---	----	---

V18 25

If 'Yes', specify:

	YES	NO
Toxoplasmosis	1	2
Cytomegalovirus	1	2
Syphilis	1	2
Herpes	1	2
Rubella	1	2
Measles	1	2
HIV	1	2
Malaria	1	2

V19 26

V20 27

V21 28

V22 29

V23 30

V24 31

V25 32

V26 33

d) Craniofacial defects (Head and neck)

Yes	1	No	2
-----	---	----	---

V27 34

e) Birth weight < 1500g

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V28 35

f) Bacterial meningitis

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V29 36

g) Asphyxia

Apgar 0-4 at 1min and/or 0-6 at 5min

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V30 37

If 'Yes' specify at:

1 min	<input type="text"/>	<input type="text"/>
5 min	<input type="text"/>	<input type="text"/>

V31 38-39

V32 40-41

h) Ototoxic medications

Used for more than 5 days (e.g. gentamycin, tobramycin, kanamycin, streptomycin, aminoglycosides and loop diuretics combined with amino's)

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V33 42



i) Persistent pulmonary hypertension / persistent fetal circulation. Prolonged mechanical ventilation ≥ 5 days

Yes	1	No	2	Info unavailable	3
-----	---	----	---	------------------	---

V34		43
-----	--	----

j) Syndrome present

Yes	1	No	2
-----	---	----	---

V35		44
-----	--	----

If 'yes', specify syndrome:

V36			45 - 46
-----	--	--	---------

k) Admitted to the NICU

Yes	1	No	2
-----	---	----	---

V37		47
-----	--	----

If 'Yes', for how long?

No of days			
------------	--	--	--

V38				48-50
-----	--	--	--	-------

SECTION C ~ IMMITTANCE

a) 1000 Hz Tympanogram

i. Y – Admittance

	RIGHT		LEFT	
i.i Performed	Yes	No	Yes	No
i.ii Discernable peak	Yes	No	Yes	No
i.iii Admittance (mmho)		,		,
i.iv Pressure (daPa)				
i.v Double peak	Yes	No	Yes	No
i.vi Time taken (min)				

V39R		51	
V39L		52	
V40R		53	
V40L		54	
V41R	,		55-58
V41L	,		59-62
V42R			63-66
V42L			67-70
V43R			71
V43L			72
V44R			73-74
V44L			75-76

ii. B – Susceptance

	RIGHT		LEFT	
ii.i Performed	Yes	No	Yes	No
ii.ii Admittance (mmho)		,		,
ii.iii Pressure (daPa)				

V45R		77	
V45L		78	
V46R	,		79-82
V46L	,		83-86
V47R			87-90
V47L			91-94



iii. G– Conductance

	RIGHT		LEFT	
	Yes	No	Yes	No
iii.i Performed				
iii.ii Admittance (mmho)				
iii.iii Pressure (daPa)				

V48R		95
V48L		96
V49R		97-100
V49L		101-104
V50R		105-108
V50L		109-112

b) 1000 Hz Probe Tone Reflex

	RIGHT		LEFT	
	Yes	No	Yes	No
i. Performed				
ii. Threshold present				
iii. Threshold value (dB)				

V51R		113
V51L		114
V52R		115
V52L		116
V53R		117-119
V53L		120-122

SECTION D ~ HEARING SCREENING

a) First Screen

	RIGHT		LEFT	
	Pass	Refer	Pass	Refer
i. OAE				
ii. AABR				

V54R		123
V54L		124
V55R		125
V55L		126

iii. Time taken:

iii.i OAE

		min
--	--	-----

V56		127-128
-----	--	---------

iii.ii AABR

		min
--	--	-----

V57		129-130
-----	--	---------



b) Follow-up Screen

i. Returned?

Yes	1	No	2
-----	---	----	---

V58 131

	RIGHT		LEFT	
ii. OAE	Pass	Refer	Pass	Refer
iii. AABR	Pass	Refer	Pass	Refer

V59R	<input type="text"/>	132
V59L	<input type="text"/>	133
V60R	<input type="text"/>	134
V60L	<input type="text"/>	135

SECTION E ~ DIAGNOSTIC ASSESSMENT

a) Returned?

Yes	1	No	2
-----	---	----	---

V61 136

b) Hearing loss?

None	1	Bilateral	2	Unilateral	3
------	---	-----------	---	------------	---

V62 137

c) Type of hearing loss?

S/N	1
Conductive	2
Mixed	3
AN	4

V63 138

d) Ear

Left	1	Right	2	Both	3
------	---	-------	---	------	---

V64 139

e) Degree of hearing loss?

i. Right ear

Mild (15-30 dB)	1
Moderate (31-50dB)	2
Severe (51-70dB)	3
Profound (71dB+)	4

V65 140

ii. Left ear

Mild (15-30 dB)	1
Moderate (31-50dB)	2
Severe (51-70dB)	3
Profound (71dB+)	4


V66 141



COMMENTS



APPENDIX B: Ethical Clearance



University of Pretoria

Pretoria 0002 Republic of South Africa Tel 012-420-2357
/ 012-420-2816 Fax 012-420-3517 <http://www.up.ac.za>

Department of Communication Pathology
Speech, Voice and Hearing Clinic

18 November 2003

To: Chair Research Proposal and Ethics Committee, Faculty of Humanities,
University of Pretoria

Re: M. Communication Pathology research project: Ms. S Werner (9807345)

The student participated in a larger research project initiated by Mr. DCD Swanepoel for which ethical clearance was obtained at the end of 2002 (See attached clearance letter). Ms Werner collected a sub-set of data for her research project that was included in the research proposal and ethical applications form submitted by Mr Swanepoel. The project entailed data collection at maternal child health clinics in the Hammanskraal district and ethical clearance was therefore also obtained from the North West Province Department of Health for the Moretele District (See attached clearance letter). The title registration for Ms Werner is attached for registration purposes.

Thank you,

Mr. DCD Swanepoel

DEPARTMENT COMMUNICATION PATHOLOGY